

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070620\
 Data File : VX017181.D
 Acq On : 06 Jul 2020 20:12
 Operator : JC/SP
 Sample : L3159-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 41BR-20200629

Quant Time: Jul 07 07:53:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	265589	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	441258	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	457454	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	257026	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	168648	51.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.02%	
35) Dibromofluoromethane	5.46	113	147358	52.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.58%	
50) Toluene-d8	8.70	98	539859	52.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.60%	
62) 4-Bromofluorobenzene	11.12	95	226444	56.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.24%	
Target Compounds						
4) Vinyl Chloride	1.39	62	29621	10.265	ug/l	99
16) Acetone	2.42	43	5855	4.780	ug/l #	85
17) Carbon Disulfide	2.55	76	4453	0.630	ug/l #	92
24) 1,1-Dichloroethane	3.67	63	3152	0.616	ug/l #	89
27) cis-1,2-Dichloroethene	4.57	96	25589	7.745	ug/l	89
52) Toluene	8.76	92	7828	1.023	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070620\
Data File : VX017181.D
Acq On : 06 Jul 2020 20:12
Operator : JC/SP
Sample : L3159-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
41BR-20200629

Quant Time: Jul 07 07:53:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

