

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070620\
 Data File : VX017185.D
 Acq On : 06 Jul 2020 21:43
 Operator : JC/SP
 Sample : L3159-19
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP05-20200629

Quant Time: Jul 07 08:00:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	270398	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	437319	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	461904	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	250691	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	168609	50.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.16%	
35) Dibromofluoromethane	5.46	113	146787	53.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.12%	
50) Toluene-d8	8.70	98	534703	52.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.54%	
62) 4-Bromofluorobenzene	11.12	95	224852	56.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.46%	
Target Compounds						
16) Acetone	2.42	43	10425	8.360	ug/l	96
27) cis-1,2-Dichloroethene	4.56	96	5037	1.497	ug/l	87
44) Trichloroethene	7.20	130	3038	0.601	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070620\
Data File : VX017185.D
Acq On : 06 Jul 2020 21:43
Operator : JC/SP
Sample : L3159-19
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP05-20200629

Quant Time: Jul 07 08:00:36 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

