

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070621\
 Data File : VX023136.D
 Acq On : 06 Jul 2021 10:56
 Operator : JC/MD
 Sample : VX0706MBL01
 Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0706MBL01

Quant Time: Jul 07 02:48:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	316060	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	515679	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	452516	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	195302	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	203977	48.019	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.040%
35) Dibromofluoromethane	5.391	113	164173	49.301	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.600%
50) Toluene-d8	8.653	98	609180	49.041	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.080%
62) 4-Bromofluorobenzene	11.085	95	218854	45.637	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070621\
Data File : VX023136.D
Acq On : 06 Jul 2021 10:56
Operator : JC/MD
Sample : VX0706MBL01
Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0706MBL01

Quant Time: Jul 07 02:48:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

