

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070621\
 Data File : VX023141.D
 Acq On : 06 Jul 2021 13:21
 Operator : JC/MD
 Sample : M2907-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-02-04-09-063021

Quant Time: Jul 07 02:50:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	272241	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	445349	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	398056	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	166749	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	183134	50.051	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 100.100%		
35) Dibromofluoromethane	5.397	113	141532	49.214	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 98.420%		
50) Toluene-d8	8.653	98	527613	49.182	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 98.360%		
62) 4-Bromofluorobenzene	11.085	95	191175	46.161	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 92.320%		
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.325	96	1706	0.703	ug/l	84
16) Acetone	2.392	43	3667	2.533	ug/l	95
24) 1,1-Dichloroethane	3.617	63	156058	31.067	ug/l	96
32) 1,1,1-Trichloroethane	5.397	97	62585	13.243	ug/l	97
44) Trichloroethene	7.135	130	3093	0.946	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070621\
Data File : VX023141.D
Acq On : 06 Jul 2021 13:21
Operator : JC/MD
Sample : M2907-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-02-04-09-063021

Quant Time: Jul 07 02:50:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

