

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070621\
 Data File : VX023145.D
 Acq On : 06 Jul 2021 14:55
 Operator : JC/MD
 Sample : M2923-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-01-070121

Quant Time: Jul 07 02:52:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	276475	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	455087	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	397267	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	166278	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	189303	50.945	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.900%
35) Dibromofluoromethane	5.397	113	142790	48.589	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.180%
50) Toluene-d8	8.653	98	535504	48.850	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.700%
62) 4-Bromofluorobenzene	11.085	95	193544	45.733	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.460%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	21954	14.934	ug/l	97
20) Methylene Chloride	2.794	84	4738	1.497	ug/l	93
25) 2-Butanone	4.580	43	6535	2.948	ug/l #	88
29) Tetrahydrofuran	5.038	42	2555	1.793	ug/l #	57

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070621\
Data File : VX023145.D
Acq On : 06 Jul 2021 14:55
Operator : JC/MD
Sample : M2923-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-01-070121

Quant Time: Jul 07 02:52:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

