

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070621\
 Data File : VX023150.D
 Acq On : 06 Jul 2021 16:53
 Operator : JC/MD
 Sample : M2923-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-05-08-13-070121

Quant Time: Jul 07 02:53:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	277176	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	451204	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	400245	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	168091	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	191376	51.373	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.740%
35) Dibromofluoromethane	5.397	113	143913	49.392	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.780%
50) Toluene-d8	8.653	98	534979	49.222	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.440%
62) 4-Bromofluorobenzene	11.085	95	193520	46.121	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.240%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	7363	2.797	ug/l	98
16) Acetone	2.386	43	4324	2.934	ug/l	94
19) Methyl tert-butyl Ether	3.123	73	44087	4.771	ug/l	96
20) Methylene Chloride	2.800	84	1854	0.584	ug/l	95
27) cis-1,2-Dichloroethene	4.495	96	22802	7.055	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070621\
Data File : VX023150.D
Acq On : 06 Jul 2021 16:53
Operator : JC/MD
Sample : M2923-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-05-08-13-070121

Quant Time: Jul 07 02:53:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

