

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070621\
 Data File : VX023153.D
 Acq On : 06 Jul 2021 18:03
 Operator : JC/MD
 Sample : M2923-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-12A-070121-FD

Quant Time: Jul 07 02:54:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	258032	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	430706	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	374425	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	148713	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.971	65	182294	52.565	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.140%
35) Dibromofluoromethane	5.397	113	136278	48.998	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.000%
50) Toluene-d8	8.653	98	506303	48.800	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.600%
62) 4-Bromofluorobenzene	11.085	95	178708	44.618	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.240%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	4430	3.229	ug/l	94
27) cis-1,2-Dichloroethene	4.501	96	10019	3.330	ug/l	92
44) Trichloroethene	7.135	130	56574	17.890	ug/l	97
64) Tetrachloroethene	9.275	164	966	0.323	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070621\
Data File : VX023153.D
Acq On : 06 Jul 2021 18:03
Operator : JC/MD
Sample : M2923-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-12A-070121-FD

Quant Time: Jul 07 02:54:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

