

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070621\
 Data File : VX023155.D
 Acq On : 06 Jul 2021 18:50
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jul 07 02:55:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	260116	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	432016	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	382431	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	158001	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	178750	51.131	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.260%
35) Dibromofluoromethane	5.397	113	133980	48.025	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.060%
50) Toluene-d8	8.653	98	512380	49.236	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.480%
62) 4-Bromofluorobenzene	11.085	95	186469	46.414	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070621\
Data File : VX023155.D
Acq On : 06 Jul 2021 18:50
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jul 07 02:55:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

