

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070621\
 Data File : VX023156.D
 Acq On : 06 Jul 2021 19:14
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 07 02:55:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	130	0.00
2 T	Dichlorodifluoromethane	50.000	51.330	-2.7	124	0.00
3 P	Chloromethane	50.000	47.878	4.2	124	0.00
4 C	Vinyl Chloride	50.000	49.330	1.3#	123	0.00
5 T	Bromomethane	50.000	42.596	14.8	110	0.00
6 T	Chloroethane	50.000	46.831	6.3	117	0.00
7 T	Trichlorofluoromethane	50.000	48.297	3.4	120	0.00
8 T	Diethyl Ether	50.000	47.869	4.3	123	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	49.276	1.4	126	0.00
10 T	Methyl Iodide	50.000	46.250	7.5	118	0.00
11 T	Tert butyl alcohol	250.000	245.091	2.0	134	0.00
12 CM	1,1-Dichloroethene	50.000	49.476	1.0#	124	0.00
13 T	Acrolein	250.000	182.061	27.2#	95	0.00
14 T	Allyl chloride	50.000	50.660	-1.3	127	0.00
15 T	Acrylonitrile	250.000	247.880	0.8	125	0.00
16 T	Acetone	250.000	233.604	6.6	126	0.00
17 T	Carbon Disulfide	50.000	48.908	2.2	128	0.00
18 T	Methyl Acetate	50.000	49.005	2.0	125	0.00
19 T	Methyl tert-butyl Ether	50.000	51.137	-2.3	129	0.00
20 T	Methylene Chloride	50.000	46.211	7.6	129	0.00
21 T	trans-1,2-Dichloroethene	50.000	48.610	2.8	125	0.00
22 T	Diisopropyl ether	50.000	49.823	0.4	126	0.00
23 T	Vinyl Acetate	250.000	255.627	-2.3	125	0.00
24 P	1,1-Dichloroethane	50.000	50.095	-0.2	128	0.00
25 T	2-Butanone	250.000	243.062	2.8	123	0.00
26 T	2,2-Dichloropropane	50.000	46.965	6.1	118	0.00
27 T	cis-1,2-Dichloroethene	50.000	49.424	1.2	126	0.00
28 T	Bromochloromethane	50.000	50.693	-1.4	133	0.00
29 T	Tetrahydrofuran	250.000	241.437	3.4	122	0.00
30 C	Chloroform	50.000	50.616	-1.2#	129	0.00
31 T	Cyclohexane	50.000	47.638	4.7	125	0.00
32 T	1,1,1-Trichloroethane	50.000	51.857	-3.7	130	0.00
33 S	1,2-Dichloroethane-d4	50.000	52.775	-5.5	142	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	133	0.00
35 S	Dibromofluoromethane	50.000	51.545	-3.1	144	0.00
36 T	1,1-Dichloropropene	50.000	47.338	5.3	125	0.00
37 T	Ethyl Acetate	50.000	47.686	4.6	123	0.00
38 T	Carbon Tetrachloride	50.000	51.802	-3.6	131	0.00
39 T	Methylcyclohexane	50.000	47.789	4.4	127	0.00
40 TM	Benzene	50.000	48.499	3.0	126	0.00
41 T	Methacrylonitrile	50.000	47.618	4.8	125	0.00
42 TM	1,2-Dichloroethane	50.000	49.451	1.1	126	0.00
43 T	Isopropyl Acetate	50.000	48.401	3.2	125	0.00
44 TM	Trichloroethene	50.000	47.534	4.9	124	0.00
45 C	1,2-Dichloropropane	50.000	48.787	2.4#	127	0.00
46 T	Dibromomethane	50.000	49.316	1.4	127	0.00
47 T	Bromodichloromethane	50.000	51.944	-3.9	130	0.00
48 T	Methyl methacrylate	50.000	48.560	2.9	125	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070621\
 Data File : VX023156.D
 Acq On : 06 Jul 2021 19:14
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 07 02:55:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	975.036	2.5	125	0.00
50 S	Toluene-d8	50.000	51.320	-2.6	140	0.00
51 T	4-Methyl-2-Pentanone	250.000	243.709	2.5	123	0.00
52 CM	Toluene	50.000	48.895	2.2#	126	0.00
53 T	t-1,3-Dichloropropene	50.000	45.435	9.1	127	0.00
54 T	cis-1,3-Dichloropropene	50.000	46.048	7.9	127	0.00
55 T	1,1,2-Trichloroethane	50.000	49.321	1.4	127	0.00
56 T	Ethyl methacrylate	50.000	50.778	-1.6	129	0.00
57 T	1,3-Dichloropropane	50.000	48.553	2.9	126	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	230.585	7.8	122	0.00
59 T	2-Hexanone	250.000	243.797	2.5	122	0.00
60 T	Dibromochloromethane	50.000	45.487	9.0	129	0.00
61 T	1,2-Dibromoethane	50.000	50.978	-2.0	127	0.00
62 S	4-Bromofluorobenzene	50.000	52.491	-5.0	142	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	132	0.00
64 T	Tetrachloroethene	50.000	46.295	7.4	121	0.00
65 PM	Chlorobenzene	50.000	48.230	3.5	125	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	51.766	-3.5	128	0.00
67 C	Ethyl Benzene	50.000	49.351	1.3#	126	0.00
68 T	m/p-Xylenes	100.000	99.437	0.6	126	0.00
69 T	o-Xylene	50.000	48.868	2.3	125	0.00
70 T	Styrene	50.000	50.441	-0.9	125	0.00
71 P	Bromoform	50.000	45.163	9.7	131	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	129	0.00
73 T	Isopropylbenzene	50.000	49.753	0.5	125	0.00
74 T	N-ethyl acetate	50.000	48.902	2.2	122	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	48.484	3.0	126	0.00
76 T	1,2,3-Trichloropropane	50.000	50.784	-1.6	126	0.00
77 T	Bromobenzene	50.000	48.485	3.0	124	0.00
78 T	n-propylbenzene	50.000	49.383	1.2	123	0.00
79 T	2-Chlorotoluene	50.000	49.102	1.8	125	0.00
80 T	1,3,5-Trimethylbenzene	50.000	49.622	0.8	124	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	46.803	6.4	126	0.00
82 T	4-Chlorotoluene	50.000	49.900	0.2	124	0.00
83 T	tert-Butylbenzene	50.000	50.349	-0.7	125	0.00
84 T	1,2,4-Trimethylbenzene	50.000	50.220	-0.4	124	0.00
85 T	sec-Butylbenzene	50.000	49.767	0.5	123	0.00
86 T	p-Isopropyltoluene	50.000	50.342	-0.7	124	0.00
87 T	1,3-Dichlorobenzene	50.000	48.067	3.9	121	0.00
88 T	1,4-Dichlorobenzene	50.000	48.151	3.7	123	0.00
89 T	n-Butylbenzene	50.000	49.811	0.4	122	0.00
90 T	Hexachloroethane	50.000	47.652	4.7	133	0.00
91 T	1,2-Dichlorobenzene	50.000	48.120	3.8	122	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	46.580	6.8	127	0.00
93 T	1,2,4-Trichlorobenzene	50.000	49.165	1.7	124	0.00
94 T	Hexachlorobutadiene	50.000	48.252	3.5	123	0.00
95 T	Naphthalene	50.000	47.048	5.9	124	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070621\
Data File : VX023156.D
Acq On : 06 Jul 2021 19:14
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jul 07 02:55:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	50.375	-0.8	124	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6