

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX070621\  
 Data File : VX023138.D  
 Acq On : 06 Jul 2021 11:45  
 Operator : JC/MD  
 Sample : VX0706WBS01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0706WBS01

Manual Integrations  
 APPROVED

MMDadoda  
 7/7/2021 6:27:59 PM

Quant Time: Jul 07 02:49:23 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X070221W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Jul 02 14:29:42 2021  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response   | Conc    | Units  | Dev(Min) |
|------------------------------|----------------|------|------------|---------|--------|----------|
| -----                        |                |      |            |         |        |          |
| Internal Standards           |                |      |            |         |        |          |
| 1) Pentafluorobenzene        | 5.562          | 168  | 424738     | 50.000  | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 6.769          | 114  | 665884     | 50.000  | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 10.055         | 117  | 600388     | 50.000  | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 12.024         | 152  | 288954     | 50.000  | ug/l   | 0.00     |
| System Monitoring Compounds  |                |      |            |         |        |          |
| 33) 1,2-Dichloroethane-d4    | 5.965          | 65   | 269414     | 47.195  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 78 - 117 |      | Recovery = | 94.400% |        |          |
| 35) Dibromofluoromethane     | 5.391          | 113  | 206373     | 47.994  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 75 - 124 |      | Recovery = | 95.980% |        |          |
| 50) Toluene-d8               | 8.653          | 98   | 764927     | 47.689  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 92 - 112 |      | Recovery = | 95.380% |        |          |
| 62) 4-Bromofluorobenzene     | 11.086         | 95   | 293737     | 47.436  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 83 - 123 |      | Recovery = | 94.880% |        |          |
| Target Compounds             |                |      |            |         |        |          |
|                              |                |      |            |         | Qvalue |          |
| 2) Dichlorodifluoromethane   | 1.173          | 85   | 81314      | 21.129  | ug/l   | 96       |
| 3) Chloromethane             | 1.295          | 50   | 73205      | 19.226  | ug/l   | 99       |
| 4) Vinyl Chloride            | 1.374          | 62   | 78592      | 19.484  | ug/l   | 99       |
| 5) Bromomethane              | 1.599          | 94   | 46774      | 19.455  | ug/l   | 93       |
| 6) Chloroethane              | 1.685          | 64   | 48350      | 18.331  | ug/l   | 98       |
| 7) Trichlorofluoromethane    | 1.886          | 101  | 133085     | 18.985  | ug/l   | 97       |
| 8) Diethyl Ether             | 2.142          | 74   | 49324      | 19.064  | ug/l   | 100      |
| 9) 1,1,2-Trichlorotrifluo... | 2.331          | 101  | 84498      | 20.598  | ug/l   | 98       |
| 10) Methyl Iodide            | 2.453          | 142  | 85776      | 17.028  | ug/l   | 94       |
| 11) Tert butyl alcohol       | 2.971          | 59   | 118021     | 98.276  | ug/l   | 99       |
| 12) 1,1-Dichloroethene       | 2.319          | 96   | 75478      | 19.935  | ug/l   | 94       |
| 13) Acrolein                 | 2.240          | 56   | 63726      | 79.688  | ug/l   | 100      |
| 14) Allyl chloride           | 2.666          | 41   | 135316     | 20.567  | ug/l   | 93       |
| 15) Acrylonitrile            | 3.075          | 53   | 228099     | 97.682  | ug/l   | 98       |
| 16) Acetone                  | 2.386          | 43   | 213752     | 94.646  | ug/l   | 89       |
| 17) Carbon Disulfide         | 2.514          | 76   | 165352     | 20.707  | ug/l   | 99       |
| 18) Methyl Acetate           | 2.709          | 43   | 104889     | 19.133  | ug/l   | 93       |
| 19) Methyl tert-butyl Ether  | 3.124          | 73   | 289498     | 20.444  | ug/l   | 98       |
| 20) Methylene Chloride       | 2.794          | 84   | 87989      | 18.096  | ug/l   | 95       |
| 21) trans-1,2-Dichloroethene | 3.093          | 96   | 80545      | 19.698  | ug/l   | 93       |
| 22) Diisopropyl ether        | 3.770          | 45   | 271074     | 19.543  | ug/l   | 96       |
| 23) Vinyl Acetate            | 3.733          | 43   | 1179173    | 100.446 | ug/l   | 95       |
| 24) 1,1-Dichloroethane       | 3.617          | 63   | 154531     | 19.718  | ug/l   | 98       |
| 25) 2-Butanone               | 4.568          | 43   | 325738     | 95.647  | ug/l   | 95       |
| 26) 2,2-Dichloropropane      | 4.489          | 77   | 139532     | 21.212  | ug/l   | 99       |
| 27) cis-1,2-Dichloroethene   | 4.495          | 96   | 94731      | 19.127  | ug/l   | 95       |
| 28) Bromochloromethane       | 4.904          | 49   | 70966      | 19.087  | ug/l   | 99       |
| 29) Tetrahydrofuran          | 5.020          | 42   | 208156     | 95.084  | ug/l   | 93       |
| 30) Chloroform               | 5.105          | 83   | 164960     | 19.485  | ug/l   | 97       |
| 31) Cyclohexane              | 5.477          | 56   | 133993     | 19.604  | ug/l   | 95       |
| 32) 1,1,1-Trichloroethane    | 5.391          | 97   | 149147     | 20.228  | ug/l   | 99       |
| 36) 1,1-Dichloropropene      | 5.696          | 75   | 114847     | 19.347  | ug/l   | 99       |
| 37) Ethyl Acetate            | 4.727          | 43   | 125281     | 19.346  | ug/l   | 96       |
| 38) Carbon Tetrachloride     | 5.684          | 117  | 129719     | 20.935  | ug/l   | 98       |
| 39) Methylcyclohexane        | 7.385          | 83   | 141528     | 20.385  | ug/l   | 94       |
| 40) Benzene                  | 6.044          | 78   | 339096     | 19.675  | ug/l   | 98       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX070621\  
 Data File : VX023138.D  
 Acq On : 06 Jul 2021 11:45  
 Operator : JC/MD  
 Sample : VX0706WBS01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0706WBS01

Manual Integrations  
 APPROVED

MMDadoda  
 7/7/2021 6:27:59 PM

Quant Time: Jul 07 02:49:23 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X070221W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Jul 02 14:29:42 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|---------|--------|----------|
| 41) Methacrylonitrile         | 4.934  | 41   | 69744    | 19.266  | ug/l   | 95       |
| 42) 1,2-Dichloroethane        | 6.093  | 62   | 134466   | 19.416  | ug/l   | 96       |
| 43) Isopropyl Acetate         | 6.349  | 43   | 205114   | 19.043  | ug/l   | 96       |
| 44) Trichloroethene           | 7.129  | 130  | 93448    | 19.113  | ug/l   | 97       |
| 45) 1,2-Dichloropropane       | 7.440  | 63   | 88631    | 19.700  | ug/l   | 95       |
| 46) Dibromomethane            | 7.586  | 93   | 62233    | 19.369  | ug/l   | 96       |
| 47) Bromodichloromethane      | 7.830  | 83   | 119510   | 20.372  | ug/l   | 99       |
| 48) Methyl methacrylate       | 7.702  | 41   | 98876    | 18.708  | ug/l   | 90       |
| 49) 1,4-Dioxane               | 7.665  | 88   | 46021    | 386.888 | ug/l   | 97       |
| 51) 4-Methyl-2-Pentanone      | 8.580  | 43   | 640214   | 95.919  | ug/l   | 92       |
| 52) Toluene                   | 8.720  | 92   | 217088   | 19.463  | ug/l   | 99       |
| 53) t-1,3-Dichloropropene     | 8.982  | 75   | 132735   | 19.655  | ug/l   | 100      |
| 54) cis-1,3-Dichloropropene   | 8.373  | 75   | 144691   | 20.097  | ug/l   | 92       |
| 55) 1,1,2-Trichloroethane     | 9.159  | 97   | 91341    | 19.501  | ug/l   | 96       |
| 56) Ethyl methacrylate        | 9.122  | 69   | 143467   | 19.822  | ug/l # | 89       |
| 57) 1,3-Dichloropropane       | 9.311  | 76   | 148500   | 19.182  | ug/l   | 100      |
| 58) 2-Chloroethyl Vinyl ether | 8.245  | 63   | 348073   | 91.396  | ug/l   | 99       |
| 59) 2-Hexanone                | 9.433  | 43   | 484070   | 94.523  | ug/l   | 90       |
| 60) Dibromochloromethane      | 9.525  | 129  | 90355    | 19.313  | ug/l   | 100      |
| 61) 1,2-Dibromoethane         | 9.616  | 107  | 95091    | 19.781  | ug/l   | 99       |
| 64) Tetrachloroethene         | 9.275  | 164  | 93158    | 19.454  | ug/l   | 93       |
| 65) Chlorobenzene             | 10.086 | 112  | 236520   | 19.929  | ug/l   | 99       |
| 66) 1,1,1,2-Tetrachloroethane | 10.165 | 131  | 90442    | 21.076  | ug/l   | 100      |
| 67) Ethyl Benzene             | 10.195 | 91   | 423079   | 20.165  | ug/l   | 99       |
| 68) m/p-Xylenes               | 10.305 | 106  | 324576   | 40.521  | ug/l   | 97       |
| 69) o-Xylene                  | 10.647 | 106  | 161444   | 19.999  | ug/l   | 100      |
| 70) Styrene                   | 10.659 | 104  | 265677   | 20.147  | ug/l   | 98       |
| 71) Bromoform                 | 10.805 | 173  | 64416    | 20.284  | ug/l # | 98       |
| 73) Isopropylbenzene          | 10.964 | 105  | 430331   | 21.140  | ug/l   | 99       |
| 74) N-amyl acetate            | 10.848 | 43   | 174466   | 20.198  | ug/l   | 93       |
| 75) 1,1,2,2-Tetrachloroethane | 11.214 | 83   | 129781   | 20.067  | ug/l   | 97       |
| 76) 1,2,3-Trichloropropane    | 11.244 | 75   | 125005m  | 21.498  | ug/l   |          |
| 77) Bromobenzene              | 11.201 | 156  | 103577   | 20.380  | ug/l   | 97       |
| 78) n-propylbenzene           | 11.305 | 91   | 492693   | 20.931  | ug/l   | 99       |
| 79) 2-Chlorotoluene           | 11.366 | 91   | 301208   | 20.705  | ug/l   | 99       |
| 80) 1,3,5-Trimethylbenzene    | 11.457 | 105  | 365751   | 20.868  | ug/l   | 100      |
| 81) trans-1,4-Dichloro-2-b... | 11.025 | 75   | 41572    | 22.829  | ug/l   | 93       |
| 82) 4-Chlorotoluene           | 11.457 | 91   | 345376   | 21.023  | ug/l   | 99       |
| 83) tert-Butylbenzene         | 11.720 | 119  | 352050   | 21.162  | ug/l   | 94       |
| 84) 1,2,4-Trimethylbenzene    | 11.756 | 105  | 367235   | 21.202  | ug/l   | 98       |
| 85) sec-Butylbenzene          | 11.896 | 105  | 454267   | 21.222  | ug/l   | 100      |
| 86) p-Isopropyltoluene        | 12.012 | 119  | 376428   | 20.963  | ug/l   | 97       |
| 87) 1,3-Dichlorobenzene       | 11.976 | 146  | 191257   | 20.040  | ug/l   | 99       |
| 88) 1,4-Dichlorobenzene       | 12.043 | 146  | 192384   | 20.020  | ug/l   | 97       |
| 89) n-Butylbenzene            | 12.335 | 91   | 324932   | 20.657  | ug/l   | 98       |
| 90) Hexachloroethane          | 12.543 | 117  | 58620    | 20.936  | ug/l   | 96       |
| 91) 1,2-Dichlorobenzene       | 12.341 | 146  | 192151   | 20.239  | ug/l   | 99       |
| 92) 1,2-Dibromo-3-Chloropr... | 12.945 | 75   | 29902    | 20.449  | ug/l   | 91       |
| 93) 1,2,4-Trichlorobenzene    | 13.591 | 180  | 118082   | 20.191  | ug/l   | 98       |
| 94) Hexachlorobutadiene       | 13.725 | 225  | 52917    | 21.370  | ug/l   | 98       |
| 95) Naphthalene               | 13.780 | 128  | 405452   | 19.591  | ug/l   | 100      |
| 96) 1,2,3-Trichlorobenzene    | 13.963 | 180  | 118911   | 20.217  | ug/l   | 97       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX070621\  
Data File : VX023138.D  
Acq On : 06 Jul 2021 11:45  
Operator : JC/MD  
Sample : VX0706WBS01  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 5 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01

Manual Integrations  
APPROVED

MMDadoda  
7/7/2021 6:27:59 PM

Quant Time: Jul 07 02:49:23 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X070221W.M  
Quant Title : SW846 8260  
QLast Update : Fri Jul 02 14:29:42 2021  
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX070621\  
 Data File : VX023138.D  
 Acq On : 06 Jul 2021 11:45  
 Operator : JC/MD  
 Sample : VX0706WBS01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 5 Sample Multiplier: 1


Instrument :  
 MSVOA\_X  
 Client Sampled :  
 VX0706WBS01

Manual Integrations  
 APPROVED


MMDadoda  
 7/7/2021 6:27:59 PM

Quant Time: Jul 07 02:49:23 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X070221W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Jul 02 14:29:42 2021  
 Response via : Initial Calibration







#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 5.562 min Scan# 734  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45




Tgt Ion:168 Resp: 424738  
Ion Ratio Lower Upper  
168 100  
99 60.8 45.8 68.6




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01

Manual Integrations  
APPROVED



MMDadoda  
7/7/2021 6:27:59 PM




#2  
Dichlorodifluoromethane  
Concen: 21.129 ug/l  
RT: 1.173 min Scan# 14  
Delta R.T. 0.001 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45




Tgt Ion: 85 Resp: 81314  
Ion Ratio Lower Upper  
85 100  
87 30.6 16.4 49.2





#3  
Chloromethane  
Concen: 19.226 ug/l  
RT: 1.295 min Scan# 34  
Delta R.T. 0.001 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01




Tgt Ion: 50 Resp: 73205  
Ion Ratio Lower Upper  
50 100  
52 32.2 26.4 39.6

Manual Integrations  
APPROVED



MMDadoda  
7/7/2021 6:27:59 PM




#4  
Vinyl Chloride  
Concen: 19.484 ug/l  
RT: 1.374 min Scan# 47  
Delta R.T. -0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45




Tgt Ion: 62 Resp: 78592  
Ion Ratio Lower Upper  
62 100  
64 31.4 25.7 38.5







#5  
Bromomethane  
Concen: 19.455 ug/l  
RT: 1.599 min Scan# 84  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45




Tgt Ion: 94 Resp: 46774  
Ion Ratio Lower Upper  
94 100  
96 89.1 76.4 114.6




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01

Manual Integrations  
APPROVED



MMDadoda  
7/7/2021 6:27:59 PM




#6  
Chloroethane  
Concen: 18.331 ug/l  
RT: 1.685 min Scan# 98  
Delta R.T. 0.006 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45



Tgt Ion: 64 Resp: 48350  
Ion Ratio Lower Upper  
64 100  
66 32.7 25.4 38.0









#7  
Trichlorofluoromethane  
Concen: 18.985 ug/l  
RT: 1.886 min Scan# 131  
Delta R.T. -0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 101 Resp: 133085  
Ion Ratio Lower Upper  
101 100  
103 64.4 53.2 79.8

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01




Manual Integrations  
APPROVED

MMDadoda  
7/7/2021 6:27:59 PM




#8  
Diethyl Ether  
Concen: 19.064 ug/l  
RT: 2.142 min Scan# 173  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 74 Resp: 49324  
Ion Ratio Lower Upper  
74 100  
45 89.1 44.5 133.5











#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 20.598 ug/l  
RT: 2.331 min Scan# 204  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion:101 Resp: 84498  
Ion Ratio Lower Upper  
101 100  
85 44.2 34.9 52.3  
151 79.3 61.4 92.2

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01




Manual Integrations  
APPROVED


MMDadoda  
7/7/2021 6:27:59 PM



#10  
Methyl Iodide  
Concen: 17.028 ug/l  
RT: 2.453 min Scan# 224  
Delta R.T. -0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion:142 Resp: 85776  
Ion Ratio Lower Upper  
142 100  
127 46.4 33.3 49.9  
141 14.5 11.4 17.0









#11  
Tert butyl alcohol  
Concen: 98.276 ug/l  
RT: 2.971 min Scan# 309  
Delta R.T. -0.012 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 59 Resp: 118021  
Ion Ratio Lower Upper  
59 100  
57 10.3 7.8 11.8

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01




Manual Integrations  
APPROVED


MMDadoda  
7/7/2021 6:27:59 PM



#12  
1,1-Dichloroethene  
Concen: 19.935 ug/l  
RT: 2.319 min Scan# 202  
Delta R.T. -0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 96 Resp: 75478  
Ion Ratio Lower Upper  
96 100  
61 165.2 124.6 187.0  
98 61.5 50.6 76.0






#13  
Acrolein  
Concen: 79.688 ug/l  
RT: 2.240 min Scan# 189  
Delta R.T. -0.005 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01

Manual Integrations  
APPROVED

MMDadoda  
7/7/2021 6:27:59 PM







Tgt Ion: 56 Resp: 63726  
Ion Ratio Lower Upper  
56 100  
55 70.6 56.4 84.6



#14  
Allyl chloride  
Concen: 20.567 ug/l  
RT: 2.666 min Scan# 259  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 41 Resp: 135316  
Ion Ratio Lower Upper  
41 100  
39 69.0 51.6 77.4  
76 33.3 30.5 45.7









#15  
Acrylonitrile  
Concen: 97.682 ug/l  
RT: 3.075 min Scan# 326  
Delta R.T. -0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 53 Resp: 228099  
Ion Ratio Lower Upper  
53 100  
52 82.4 65.3 97.9  
51 37.4 28.2 42.2

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01




Manual Integrations  
APPROVED


MMDadoda  
7/7/2021 6:27:59 PM



#16  
Acetone  
Concen: 94.646 ug/l  
RT: 2.386 min Scan# 213  
Delta R.T. -0.006 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 43 Resp: 213752  
Ion Ratio Lower Upper  
43 100  
58 28.4 27.8 41.8









#17  
Carbon Disulfide  
Concen: 20.707 ug/l  
RT: 2.514 min Scan# 234  
Delta R.T. -0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 76 Resp: 165352  
Ion Ratio Lower Upper  
76 100  
78 8.5 7.2 10.8

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01




Manual Integrations  
APPROVED


MMDadoda  
7/7/2021 6:27:59 PM



#18  
Methyl Acetate  
Concen: 19.133 ug/l  
RT: 2.709 min Scan# 266  
Delta R.T. -0.006 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 43 Resp: 104889  
Ion Ratio Lower Upper  
43 100  
74 23.9 21.9 32.9









#19  
Methyl tert-butyl Ether  
Concen: 20.444 ug/l  
RT: 3.124 min Scan# 334  
Delta R.T. 0.001 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 73 Resp: 289498  
Ion Ratio Lower Upper  
73 100  
57 20.7 17.1 25.7

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01




Manual Integrations  
APPROVED


MMDadoda  
7/7/2021 6:27:59 PM



#20  
Methylene Chloride  
Concen: 18.096 ug/l  
RT: 2.794 min Scan# 280  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 84 Resp: 87989  
Ion Ratio Lower Upper  
84 100  
49 118.2 88.6 132.8  
51 35.5 27.0 40.6  
86 63.7 50.3 75.5









#21  
trans-1,2-Dichloroethene  
Concen: 19.698 ug/l  
RT: 3.093 min Scan# 329  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 96 Resp: 80545  
Ion Ratio Lower Upper  
96 100  
61 144.2 107.7 161.5  
98 60.1 50.4 75.6

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01




Manual Integrations  
APPROVED


MMDadoda  
7/7/2021 6:27:59 PM



#22  
Diisopropyl ether  
Concen: 19.543 ug/l  
RT: 3.770 min Scan# 440  
Delta R.T. -0.006 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 45 Resp: 271074  
Ion Ratio Lower Upper  
45 100  
43 77.4 58.0 87.0  
87 31.2 25.4 38.0  
59 11.7 10.2 15.2






#23  
 Vinyl Acetate  
 Concen: 100.446 ug/l  
 RT: 3.733 min Scan# 434  
 Delta R.T. 0.000 min  
 Lab File: VX023138.D  
 Acq: 06 Jul 2021 11:45

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0706WBS01




Manual Integrations  
 APPROVED

MMDadoda  
 7/7/2021 6:27:59 PM



Tgt Ion: 43 Resp: 1179173  




| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 43  | 100   |       |       |
| 86  | 10.6  | 9.9   | 14.9  |




#24  
 1,1-Dichloroethane  
 Concen: 19.718 ug/l  
 RT: 3.617 min Scan# 415  
 Delta R.T. 0.000 min  
 Lab File: VX023138.D  
 Acq: 06 Jul 2021 11:45

Tgt Ion: 63 Resp: 154531  

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 63  | 100   |       |       |
| 98  | 6.5   | 3.8   | 11.2  |
| 100 | 4.4   | 2.5   | 7.5   |











#25  
2-Butanone  
Concen: 95.647 ug/l  
RT: 4.568 min Scan# 571  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 43 Resp: 325738  
Ion Ratio Lower Upper  
43 100  
72 26.7 23.7 35.5

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01




Manual Integrations  
APPROVED


MMDadoda  
7/7/2021 6:27:59 PM



#26  
2,2-Dichloropropane  
Concen: 21.212 ug/l  
RT: 4.489 min Scan# 558  
Delta R.T. 0.006 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45


Tgt Ion: 77 Resp: 139532  
Ion Ratio Lower Upper  
77 100  
97 22.7 11.7 35.1





#27  
cis-1,2-Dichloroethene  
Concen: 19.127 ug/l  
RT: 4.495 min Scan# 559  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01




Tgt Ion: 96 Resp: 94731  
Ion Ratio Lower Upper  
96 100  
61 150.6 0.0 285.6  
98 65.9 0.0 129.6

Manual Integrations  
APPROVED



MMDadoda  
7/7/2021 6:27:59 PM




#28  
Bromochloromethane  
Concen: 19.087 ug/l  
RT: 4.904 min Scan# 626  
Delta R.T. -0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45



Tgt Ion: 49 Resp: 70966  
Ion Ratio Lower Upper  
49 100  
129 2.1 0.0 4.0  
130 84.2 68.2 102.2





#29

Tetrahydrofuran

Concen: 95.084 ug/l

RT: 5.020 min Scan# 645

Delta R.T. 0.001 min

Lab File: VX023138.D





Acq: 06 Jul 2021 11:45

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 42    | Resp: | 208156 |
| Ion      | Ratio | Lower | Upper  |
| 42       | 100   |       |        |
| 72       | 46.9  | 41.7  | 62.5   |
| 71       | 44.2  | 38.4  | 57.6   |

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01

Manual Integrations  
APPROVED

MMDadoda  
7/7/2021 6:27:59 PM



#30

Chloroform

Concen: 19.485 ug/l




RT: 5.105 min Scan# 659


Delta R.T. -0.000 min

Lab File: VX023138.D

Acq: 06 Jul 2021 11:45


|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 83    | Resp: | 164960 |
| Ion      | Ratio | Lower | Upper  |
| 83       | 100   |       |        |
| 85       | 67.2  | 52.0  | 78.0   |








#31  
Cyclohexane  
Concen: 19.604 ug/l  
RT: 5.477 min Scan# 720  
Delta R.T. -0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45


Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01





Tgt Ion: 56 Resp: 133993  
Ion Ratio Lower Upper  
56 100  
69 34.7 25.3 37.9  
84 84.3 71.4 107.0




#32  
1,1,1-Trichloroethane  
Concen: 20.228 ug/l  
RT: 5.391 min Scan# 706  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45



Tgt Ion: 97 Resp: 149147  
Ion Ratio Lower Upper  
97 100  
99 65.0 51.5 77.3  
61 43.5 34.2 51.4





#33





1,2-Dichloroethane-d4  
Concen: 47.195 ug/l  
RT: 5.965 min Scan# 800  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 65 Resp: 269414  
Ion Ratio Lower Upper  
65 100  
67 51.4 0.0 105.4

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01

Manual Integrations  
APPROVED




MMDadoda  
7/7/2021 6:27:59 PM




#34

1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 6.769 min Scan# 932  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 114 Resp: 665884  
Ion Ratio Lower Upper  
114 100  
63 20.1 0.0 40.8  
88 16.4 0.0 32.8









#35  
Dibromofluoromethane  
Concen: 47.994 ug/l  
RT: 5.391 min Scan# 706  
Delta R.T. -0.006 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 113 Resp: 206373  
Ion Ratio Lower Upper  
113 100  
111 102.2 81.9 122.9  
192 19.2 16.7 25.1

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01




Manual Integrations  
APPROVED


MMDadoda  
7/7/2021 6:27:59 PM



#36  
1,1-Dichloropropene  
Concen: 19.347 ug/l  
RT: 5.696 min Scan# 756  
Delta R.T. -0.006 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45


Tgt Ion: 75 Resp: 114847  
Ion Ratio Lower Upper  
75 100  
110 36.8 18.1 54.4  
77 31.1 25.3 37.9





#37  
Ethyl Acetate  
Concen: 19.346 ug/l  
RT: 4.727 min Scan# 597  
Delta R.T. -0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01




Tgt Ion: 43 Resp: 125281  
Ion Ratio Lower Upper  
43 100  
61 14.5 12.8 19.2  
70 11.4 10.5 15.7

Manual Integrations  
APPROVED



MMDadoda  
7/7/2021 6:27:59 PM




#38  
Carbon Tetrachloride  
Concen: 20.935 ug/l  
RT: 5.684 min Scan# 754  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45



Tgt Ion: 117 Resp: 129719  
Ion Ratio Lower Upper  
117 100  
119 96.9 76.2 114.2  
121 29.8 24.8 37.2









#39  
Methylcyclohexane  
Concen: 20.385 ug/l  
RT: 7.385 min Scan# 1033  
Delta R.T. -0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 83 Resp: 141528  
Ion Ratio Lower Upper  
83 100  
55 79.6 58.9 88.3  
98 44.7 38.1 57.1

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01




Manual Integrations  
APPROVED

MMDadoda  
7/7/2021 6:27:59 PM




#40  
Benzene  
Concen: 19.675 ug/l  
RT: 6.044 min Scan# 813  
Delta R.T. -0.006 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45


Tgt Ion: 78 Resp: 339096  
Ion Ratio Lower Upper  
78 100  
77 22.6 18.9 28.3










#41  
Methacrylonitrile  
Concen: 19.266 ug/l  
RT: 4.934 min Scan# 631  
Delta R.T. 0.006 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45




Tgt Ion: 41 Resp: 69744

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 41  | 100   |       |       |
| 39  | 56.4  | 42.5  | 63.7  |
| 67  | 75.3  | 65.0  | 97.6  |
| 52  | 29.7  | 25.3  | 37.9  |





#42  
1,2-Dichloroethane  
Concen: 19.416 ug/l  
RT: 6.093 min Scan# 821  
Delta R.T. -0.005 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45



Tgt Ion: 62 Resp: 134466


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 62  | 100   |       |       |
| 98  | 8.5   | 0.0   | 20.2  |



Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01


Manual Integrations  
APPROVED

MMDadoda  
7/7/2021 6:27:59 PM



#43  
Isopropyl Acetate  
Concen: 19.043 ug/l  
RT: 6.349 min Scan# 863  
Delta R.T. -0.005 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01




Tgt Ion: 43 Resp: 205114  
Ion Ratio Lower Upper  
43 100  
61 20.6 18.2 27.4  
87 14.2 12.5 18.7

Manual Integrations  
APPROVED



MMDadoda  
7/7/2021 6:27:59 PM




#44  
Trichloroethene  
Concen: 19.113 ug/l  
RT: 7.129 min Scan# 991  
Delta R.T. -0.006 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45



Tgt Ion: 130 Resp: 93448  
Ion Ratio Lower Upper  
130 100  
95 101.7 0.0 198.0





#45





1,2-Dichloropropane  
Concen: 19.700 ug/l  
RT: 7.440 min Scan# 1042  
Delta R.T. -0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 63 Resp: 88631  
Ion Ratio Lower Upper  
63 100  
65 29.3 25.7 38.5

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01

Manual Integrations  
APPROVED




MMDadoda  
7/7/2021 6:27:59 PM




#46


Dibromomethane  
Concen: 19.369 ug/l  
RT: 7.586 min Scan# 1066  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 93 Resp: 62233  
Ion Ratio Lower Upper  
93 100  
95 82.2 66.8 100.2  
174 97.1 83.1 124.7







#47  
Bromodichloromethane  
Concen: 20.372 ug/l  
RT: 7.830 min Scan# 1106  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45




Tgt Ion: 83 Resp: 119510  
Ion Ratio Lower Upper  
83 100  
85 63.1 51.4 77.0  
127 9.0 6.8 10.2




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01

Manual Integrations  
APPROVED



MMDadoda  
7/7/2021 6:27:59 PM




#48  
Methyl methacrylate  
Concen: 18.708 ug/l  
RT: 7.702 min Scan# 1085  
Delta R.T. -0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45



Tgt Ion: 41 Resp: 98876  
Ion Ratio Lower Upper  
41 100  
69 83.4 76.0 114.0  
39 60.5 44.6 67.0









#49  
1,4-Dioxane  
Concen: 386.888 ug/l  
RT: 7.665 min Scan# 1079  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 88 Resp: 46021  
Ion Ratio Lower Upper  
88 100  
43 32.0 23.2 34.8  
58 70.6 55.0 82.4

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01




Manual Integrations  
APPROVED


MMDadoda  
7/7/2021 6:27:59 PM



#50  
Toluene-d8  
Concen: 47.689 ug/l  
RT: 8.653 min Scan# 1241  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 98 Resp: 764927  
Ion Ratio Lower Upper  
98 100  
100 64.9 51.9 77.9









#51  
4-Methyl-2-Pentanone  
Concen: 95.919 ug/l  
RT: 8.580 min Scan# 1229  
Delta R.T. -0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 43 Resp: 640214  
Ion Ratio Lower Upper  
43 100  
58 39.2 35.4 53.0

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01




Manual Integrations  
APPROVED


MMDadoda  
7/7/2021 6:27:59 PM



#52  
Toluene  
Concen: 19.463 ug/l  
RT: 8.720 min Scan# 1252  
Delta R.T. -0.006 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 92 Resp: 217088  
Ion Ratio Lower Upper  
92 100  
91 172.5 137.1 205.7









#53  
t-1,3-Dichloropropene  
Concen: 19.655 ug/l  
RT: 8.982 min Scan# 1295  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 75 Resp: 132735  
Ion Ratio Lower Upper  
75 100  
77 31.3 24.9 37.3

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01




Manual Integrations  
APPROVED


MMDadoda  
7/7/2021 6:27:59 PM



#54  
cis-1,3-Dichloropropene  
Concen: 20.097 ug/l  
RT: 8.373 min Scan# 1195  
Delta R.T. 0.001 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 75 Resp: 144691  
Ion Ratio Lower Upper  
75 100  
77 30.7 25.7 38.5  
39 51.3 35.1 52.7









#55  
1,1,2-Trichloroethane  
Concen: 19.501 ug/l  
RT: 9.159 min Scan# 1324  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 97    | Resp: | 91341 |
| Ion      | Ratio | Lower | Upper |
| 97       | 100   |       |       |
| 83       | 80.9  | 69.3  | 103.9 |
| 85       | 51.8  | 44.3  | 66.5  |
| 99       | 62.1  | 50.3  | 75.5  |

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01




Manual Integrations  
APPROVED

MMDadoda  
7/7/2021 6:27:59 PM




#56  
Ethyl methacrylate  
Concen: 19.822 ug/l  
RT: 9.122 min Scan# 1318  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 69    | Resp: | 143467 |
| Ion      | Ratio | Lower | Upper  |
| 69       | 100   |       |        |
| 41       | 65.7  | 46.8  | 70.2   |
| 39       | 38.5  | 24.7  | 37.1#  |











#57  
1,3-Dichloropropane  
Concen: 19.182 ug/l  
RT: 9.311 min Scan# 1349  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 76 Resp: 148500  
Ion Ratio Lower Upper  
76 100  
78 32.8 26.2 39.2

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01




Manual Integrations  
APPROVED


MMDadoda  
7/7/2021 6:27:59 PM



#58  
2-Chloroethyl Vinyl ether  
Concen: 91.396 ug/l  
RT: 8.245 min Scan# 1174  
Delta R.T. -0.005 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 63 Resp: 348073  
Ion Ratio Lower Upper  
63 100  
106 29.0 23.5 35.3









#59  
2-Hexanone  
Concen: 94.523 ug/l  
RT: 9.433 min Scan# 1369  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 43 Resp: 484070  
Ion Ratio Lower Upper  
43 100  
58 54.4 30.9 92.8

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01




Manual Integrations  
APPROVED


MMDadoda  
7/7/2021 6:27:59 PM



#60  
Dibromochloromethane  
Concen: 19.313 ug/l  
RT: 9.525 min Scan# 1384  
Delta R.T. -0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 129 Resp: 90355  
Ion Ratio Lower Upper  
129 100  
127 76.8 38.4 115.2









#61  
1,2-Dibromoethane  
Concen: 19.781 ug/l  
RT: 9.616 min Scan# 1399  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion:107 Resp: 95091  
Ion Ratio Lower Upper  
107 100  
109 94.0 75.8 113.6

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01




Manual Integrations  
APPROVED


MMDadoda  
7/7/2021 6:27:59 PM



#62  
4-Bromofluorobenzene  
Concen: 47.436 ug/l  
RT: 11.086 min Scan# 1640  
Delta R.T. 0.001 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45


Tgt Ion: 95 Resp: 293737  
Ion Ratio Lower Upper  
95 100  
174 75.4 0.0 154.6  
176 72.6 0.0 155.8





#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 10.055 min Scan# 1471  
Delta R.T. -0.006 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01




Tgt Ion:117 Resp: 600388  
Ion Ratio Lower Upper  
117 100  
82 56.2 47.4 71.2  
119 31.4 26.6 39.8

Manual Integrations  
APPROVED



MMDadoda  
7/7/2021 6:27:59 PM




#64  
Tetrachloroethene  
Concen: 19.454 ug/l  
RT: 9.275 min Scan# 1343  
Delta R.T. -0.006 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45




Tgt Ion:164 Resp: 93158  
Ion Ratio Lower Upper  
164 100  
166 125.7 103.4 155.2  
129 99.7 70.8 106.2  
131 96.8 72.5 108.7







#65  
Chlorobenzene  
Concen: 19.929 ug/l  
RT: 10.086 min Scan# 1476  
Delta R.T. -0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45




Tgt Ion:112 Resp: 236520  
Ion Ratio Lower Upper  
112 100  
114 31.9 25.8 38.6




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01

Manual Integrations  
APPROVED



MMDadoda  
7/7/2021 6:27:59 PM




#66  
1,1,1,2-Tetrachloroethane  
Concen: 21.076 ug/l  
RT: 10.165 min Scan# 1489  
Delta R.T. -0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45



Tgt Ion:131 Resp: 90442  
Ion Ratio Lower Upper  
131 100  
133 95.6 47.7 143.1  
119 64.4 32.1 96.5









#67  
Ethyl Benzene  
Concen: 20.165 ug/l  
RT: 10.195 min Scan# 1494  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 91 Resp: 423079  
Ion Ratio Lower Upper  
91 100  
106 31.6 24.6 37.0

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01




Manual Integrations  
APPROVED


MMDadoda  
7/7/2021 6:27:59 PM




#68  
m/p-Xylenes  
Concen: 40.521 ug/l  
RT: 10.305 min Scan# 1512  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 106 Resp: 324576  
Ion Ratio Lower Upper  
106 100  
91 207.4 162.8 244.2









#69  
o-Xylene  
Concen: 19.999 ug/l  
RT: 10.647 min Scan# 1568  
Delta R.T. 0.001 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45





Tgt Ion:106 Resp: 161444  
Ion Ratio Lower Upper  
106 100  
91 216.9 108.1 324.2



#70  
Styrene  
Concen: 20.147 ug/l  
RT: 10.659 min Scan# 1570  
Delta R.T. -0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45




Tgt Ion:104 Resp: 265677  
Ion Ratio Lower Upper  
104 100  
78 52.7 40.0 60.0  
103 55.0 43.4 65.0




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01

Manual Integrations  
APPROVED




MMDadoda  
7/7/2021 6:27:59 PM




#71  
Bromoform  
Concen: 20.284 ug/l  
RT: 10.805 min Scan# 1594  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45





Tgt Ion:173 Resp: 64416  
Ion Ratio Lower Upper  
173 100  
175 46.8 23.9 71.8  
254 0.0 0.1 0.1#



#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 12.024 min Scan# 1794  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45



Tgt Ion:152 Resp: 288954  
Ion Ratio Lower Upper  
152 100  
115 71.4 41.1 123.3  
150 163.6 0.0 349.0




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01

Manual Integrations  
APPROVED

MMDadoda  
7/7/2021 6:27:59 PM









#73  
Isopropylbenzene  
Concen: 21.140 ug/l  
RT: 10.964 min Scan# 1620  
Delta R.T. 0.001 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 105 Resp: 430331  
Ion Ratio Lower Upper  
105 100  
120 26.1 13.2 39.6

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01




Manual Integrations  
APPROVED


MMDadoda  
7/7/2021 6:27:59 PM



#74  
N-ethyl acetate  
Concen: 20.198 ug/l  
RT: 10.848 min Scan# 1601  
Delta R.T. -0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 43 Resp: 174466  
Ion Ratio Lower Upper  
43 100  
70 43.0 39.0 58.6  
55 25.0 21.5 32.3  
61 24.2 22.1 33.1









#75  
1,1,2,2-Tetrachloroethane  
Concen: 20.067 ug/l  
RT: 11.214 min Scan# 1661  
Delta R.T. -0.005 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 83 Resp: 129781  
Ion Ratio Lower Upper  
83 100  
131 11.2 5.1 15.2  
85 65.6 31.8 95.4

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01




Manual Integrations  
APPROVED


MMDadoda  
7/7/2021 6:27:59 PM



#76  
1,2,3-Trichloropropane  
Concen: 21.498 ug/l m  
RT: 11.244 min Scan# 1666  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 75 Resp: 125005  
Ion Ratio Lower Upper  
75 100  
77 39.7 19.1 57.5





#77  
Bromobenzene  
Concen: 20.380 ug/l  
RT: 11.201 min Scan# 1659  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45





Tgt Ion: 156 Resp: 103577  

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 156 | 100   |       |       |
| 77  | 161.6 | 82.8  | 248.6 |
| 158 | 96.0  | 49.1  | 147.4 |

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01

Manual Integrations  
APPROVED




MMDadoda  
7/7/2021 6:27:59 PM




#78  
n-propylbenzene  
Concen: 20.931 ug/l  
RT: 11.305 min Scan# 1676  
Delta R.T. -0.006 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 91 Resp: 492693  

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 91  | 100   |       |       |
| 120 | 23.3  | 11.5  | 34.5  |









#79  
2-Chlorotoluene  
Concen: 20.705 ug/l  
RT: 11.366 min Scan# 1686  
Delta R.T. -0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 91 Resp: 301208  
Ion Ratio Lower Upper  
91 100  
126 33.6 16.6 49.8

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01




Manual Integrations  
APPROVED


MMDadoda  
7/7/2021 6:27:59 PM



#80  
1,3,5-Trimethylbenzene  
Concen: 20.868 ug/l  
RT: 11.457 min Scan# 1701  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45


Tgt Ion: 105 Resp: 365751  
Ion Ratio Lower Upper  
105 100  
120 47.5 23.8 71.5





#81  
trans-1,4-Dichloro-2-butene  
Concen: 22.829 ug/l  
RT: 11.025 min Scan# 1630  
Delta R.T. 0.001 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45




Instrument :  
MSVOA\_X  
ClientSampled :  
VX0706WBS01



Tgt Ion: 75 Resp: 41572  
Ion Ratio Lower Upper  
75 100  
53 102.9 74.5 111.7  
89 47.0 37.9 56.9




Manual Integrations  
APPROVED


MMDadoda  
7/7/2021 6:27:59 PM



#82  
4-Chlorotoluene  
Concen: 21.023 ug/l  
RT: 11.457 min Scan# 1701  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45


Tgt Ion: 91 Resp: 345376  
Ion Ratio Lower Upper  
91 100  
126 29.1 14.9 44.5





#83  
tert-Butylbenzene  
Concen: 21.162 ug/l  
RT: 11.720 min Scan# 1744  
Delta R.T. 0.001 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01




Tgt Ion:119 Resp: 352050  
Ion Ratio Lower Upper  
119 100  
91 61.9 27.9 83.7  
134 23.0 11.3 33.9

Manual Integrations  
APPROVED



MMDadoda  
7/7/2021 6:27:59 PM




#84  
1,2,4-Trimethylbenzene  
Concen: 21.202 ug/l  
RT: 11.756 min Scan# 1750  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45




Tgt Ion:105 Resp: 367235  
Ion Ratio Lower Upper  
105 100  
120 43.5 22.4 67.3





#85  
 sec-Butylbenzene  
 Concen: 21.222 ug/l  
 RT: 11.896 min Scan# 1773  
 Delta R.T. 0.000 min  
 Lab File: VX023138.D  
 Acq: 06 Jul 2021 11:45




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0706WBS01




Tgt Ion:105 Resp: 454267  
 Ion Ratio Lower Upper  
 105 100  
 134 19.8 10.0 29.8

Manual Integrations  
 APPROVED



MMDadoda  
 7/7/2021 6:27:59 PM




#86  
 p-Isopropyltoluene  
 Concen: 20.963 ug/l  
 RT: 12.012 min Scan# 1792  
 Delta R.T. 0.000 min  
 Lab File: VX023138.D  
 Acq: 06 Jul 2021 11:45



Tgt Ion:119 Resp: 376428  
 Ion Ratio Lower Upper  
 119 100  
 134 26.2 12.9 38.7  
 91 25.9 11.7 35.0









#87  
1,3-Dichlorobenzene  
Concen: 20.040 ug/l  
RT: 11.976 min Scan# 1786  
Delta R.T. 0.001 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion:146 Resp: 191257  
Ion Ratio Lower Upper  
146 100  
111 42.5 20.2 60.5  
148 63.7 31.9 95.7

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01




Manual Integrations  
APPROVED

MMDadoda  
7/7/2021 6:27:59 PM




#88  
1,4-Dichlorobenzene  
Concen: 20.020 ug/l  
RT: 12.043 min Scan# 1797  
Delta R.T. 0.001 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion:146 Resp: 192384  
Ion Ratio Lower Upper  
146 100  
111 42.6 19.6 58.8  
148 63.9 31.6 95.0








#89  
n-Butylbenzene  
Concen: 20.657 ug/l  
RT: 12.335 min Scan# 1845  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45




Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01




Tgt Ion: 91 Resp: 324932  
Ion Ratio Lower Upper  
91 100  
92 53.1 27.4 82.0  
134 25.7 13.2 39.5

Manual Integrations  
APPROVED



MMDadoda  
7/7/2021 6:27:59 PM




#90  
Hexachloroethane  
Concen: 20.936 ug/l  
RT: 12.543 min Scan# 1879  
Delta R.T. 0.001 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45



Tgt Ion: 117 Resp: 58620  
Ion Ratio Lower Upper  
117 100  
201 77.0 40.4 121.2









#91  
1,2-Dichlorobenzene  
Concen: 20.239 ug/l  
RT: 12.341 min Scan# 1846  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 146 Resp: 192151  
Ion Ratio Lower Upper  
146 100  
111 42.2 20.8 62.2  
148 62.6 31.9 95.5

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01




Manual Integrations  
APPROVED


MMDadoda  
7/7/2021 6:27:59 PM




#92  
1,2-Dibromo-3-Chloropropane  
Concen: 20.449 ug/l  
RT: 12.945 min Scan# 1945  
Delta R.T. -0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45

Tgt Ion: 75 Resp: 29902  
Ion Ratio Lower Upper  
75 100  
155 82.9 45.3 135.9  
157 105.0 57.7 173.1





#93  
1,2,4-Trichlorobenzene  
Concen: 20.191 ug/l  
RT: 13.591 min Scan# 2051  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45






Tgt Ion:180 Resp: 118082  
Ion Ratio Lower Upper  
180 100  
182 95.9 47.7 143.1  
145 33.4 15.0 45.0


Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01

Manual Integrations  
APPROVED



MMDadoda  
7/7/2021 6:27:59 PM




#94  
Hexachlorobutadiene  
Concen: 21.370 ug/l  
RT: 13.725 min Scan# 2073  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45




Tgt Ion:225 Resp: 52917  
Ion Ratio Lower Upper  
225 100  
223 63.7 31.1 93.2  
227 64.5 31.6 95.0









#95  
Naphthalene  
Concen: 19.591 ug/l  
RT: 13.780 min Scan# 2082  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45





Tgt Ion:128 Resp: 405452  
Ion Ratio Lower Upper  
128 100  
127 12.8 10.3 15.5  
129 10.7 8.8 13.2



#96  
1,2,3-Trichlorobenzene  
Concen: 20.217 ug/l  
RT: 13.963 min Scan# 2112  
Delta R.T. 0.000 min  
Lab File: VX023138.D  
Acq: 06 Jul 2021 11:45



Tgt Ion:180 Resp: 118911  
Ion Ratio Lower Upper  
180 100  
182 96.8 47.6 142.9  
145 35.8 16.1 48.3



Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0706WBS01

Manual Integrations  
APPROVED

MMDadoda  
7/7/2021 6:27:59 PM