

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070621\
 Data File : VX023155.D
 Acq On : 06 Jul 2021 18:50
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jul 07 02:55:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	260116	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	432016	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	382431	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	158001	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	178750	51.131	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.260%
35) Dibromofluoromethane	5.397	113	133980	48.025	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.060%
50) Toluene-d8	8.653	98	512380	49.236	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.480%
62) 4-Bromofluorobenzene	11.085	95	186469	46.414	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.820%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070621\
Data File : VX023155.D
Acq On : 06 Jul 2021 18:50
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jul 07 02:55:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 734
Delta R.T. -0.000 min
Lab File: VX023155.D
Acq: 06 Jul 2021 18:50




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:168 Resp: 260116
Ion Ratio Lower Upper
168 100
99 61.6 45.8 68.6



#33
1,2-Dichloroethane-d4
Concen: 51.131 ug/l
RT: 5.964 min Scan# 800
Delta R.T. 0.000 min
Lab File: VX023155.D
Acq: 06 Jul 2021 18:50

Tgt Ion: 65 Resp: 178750
Ion Ratio Lower Upper
65 100
67 51.2 0.0 105.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.769 min Scan# 932
Delta R.T. -0.000 min
Lab File: VX023155.D
Acq: 06 Jul 2021 18:50




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:114 Resp: 432016
Ion Ratio Lower Upper
114 100
63 20.6 0.0 40.8
88 16.8 0.0 32.8



#35
Dibromofluoromethane
Concen: 48.025 ug/l
RT: 5.397 min Scan# 707
Delta R.T. 0.000 min
Lab File: VX023155.D
Acq: 06 Jul 2021 18:50

Tgt Ion:113 Resp: 133980
Ion Ratio Lower Upper
113 100
111 104.8 81.9 122.9
192 19.4 16.7 25.1





#50

Toluene-d8

Concen: 49.236 ug/l

RT: 8.653 min Scan# 1241

Delta R.T. -0.000 min

Lab File: VX023155.D





Acq: 06 Jul 2021 18:50

Tgt Ion: 98 Resp: 512380

Ion Ratio Lower Upper

98 100

100 63.9 51.9 77.9



#62

4-Bromofluorobenzene

Concen: 46.414 ug/l

RT: 11.085 min Scan# 1640

Delta R.T. 0.000 min

Lab File: VX023155.D

Acq: 06 Jul 2021 18:50




Tgt Ion: 95 Resp: 186469


Ion Ratio Lower Upper

95 100

174 75.4 0.0 154.6

176 71.7 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.006 min
Lab File: VX023155.D
Acq: 06 Jul 2021 18:50

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Tgt Ion:117 Resp: 382431
Ion Ratio Lower Upper
117 100
82 59.5 47.4 71.2
119 31.6 26.6 39.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX023155.D
Acq: 06 Jul 2021 18:50

Tgt Ion:152 Resp: 158001
Ion Ratio Lower Upper
152 100
115 62.2 41.1 123.3
150 157.8 0.0 349.0

