

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070622\
 Data File : VX029933.D
 Acq On : 06 Jul 2022 10:47
 Operator : JC/MD
 Sample : VX0706WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0706WBL01

Quant Time: Jul 07 02:31:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	203667	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	347052	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	301508	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	126975	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	109378	42.426	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	84.860%
35) Dibromofluoromethane	5.385	113	100612	46.389	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.780%
50) Toluene-d8	8.653	98	398974	49.035	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.060%
62) 4-Bromofluorobenzene	11.079	95	131899	45.860	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070622\
Data File : VX029933.D
Acq On : 06 Jul 2022 10:47
Operator : JC/MD
Sample : VX0706WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0706WBL01

Quant Time: Jul 07 02:31:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

