

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070622\
 Data File : VX029937.D
 Acq On : 06 Jul 2022 12:33
 Operator : JC/MD
 Sample : N3591-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GWOW-BR-03-64-80-070522

Quant Time: Jul 07 02:34:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	199478	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	334840	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	281225	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	112578	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	106584	42.211	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	84.420%		
35) Dibromofluoromethane	5.385	113	98152	46.905	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.800%		
50) Toluene-d8	8.653	98	379600	48.355	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	96.720%		
62) 4-Bromofluorobenzene	11.085	95	119392	43.025	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	86.060%		
Target Compounds						
16) Acetone	2.386	43	2728	2.208	ug/l	94
40) Benzene	6.044	78	4487	0.499	ug/l #	88
44) Trichloroethene	7.129	130	1259	0.519	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070622\
Data File : VX029937.D
Acq On : 06 Jul 2022 12:33
Operator : JC/MD
Sample : N3591-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GWOW-BR-03-64-80-070522

Quant Time: Jul 07 02:34:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

