

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070622\
 Data File : VX029938.D
 Acq On : 06 Jul 2022 12:57
 Operator : JC/MD
 Sample : N3591-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OW-01B-64-69-070522-FD

Quant Time: Jul 07 02:34:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	193219	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	326152	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	272677	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	105570	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	102152	41.766	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	83.540%		
35) Dibromofluoromethane	5.391	113	93731	45.985	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	91.980%		
50) Toluene-d8	8.653	98	373859	48.892	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	97.780%		
62) 4-Bromofluorobenzene	11.079	95	114054	42.196	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	84.400%		
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.337	101	639	0.330	ug/l #	82
16) Acetone	2.392	43	2181	1.823	ug/l	91
27) cis-1,2-Dichloroethene	4.489	96	761	0.295	ug/l	83
44) Trichloroethene	7.129	130	8793	3.725	ug/l	90
64) Tetrachloroethene	9.275	164	2275	1.383	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070622\
Data File : VX029938.D
Acq On : 06 Jul 2022 12:57
Operator : JC/MD
Sample : N3591-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OW-01B-64-69-070522-FD

Quant Time: Jul 07 02:34:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

