

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070622\
 Data File : VX029940.D
 Acq On : 06 Jul 2022 13:44
 Operator : JC/MD
 Sample : N3591-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB-01-070522

Quant Time: Jul 07 02:35:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	210733	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	360337	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	313134	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	132704	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	113981	42.729	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	85.460%
35) Dibromofluoromethane	5.391	113	104481	46.396	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.800%
50) Toluene-d8	8.653	98	415141	49.141	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.280%
62) 4-Bromofluorobenzene	11.085	95	136821	45.817	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.640%
Target Compounds						
16) Acetone	2.392	43	2225	1.705	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070622\
Data File : VX029940.D
Acq On : 06 Jul 2022 13:44
Operator : JC/MD
Sample : N3591-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB-01-070522

Quant Time: Jul 07 02:35:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

