

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070622\
 Data File : VX029942.D
 Acq On : 06 Jul 2022 14:30
 Operator : JC/MD
 Sample : N3599-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EFF0722

Quant Time: Jul 07 02:36:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	197340	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	344878	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	300041	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	124567	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	107961	43.219	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	86.440%
35) Dibromofluoromethane	5.391	113	100216	46.497	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.000%
50) Toluene-d8	8.653	98	385944	47.732	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.460%
62) 4-Bromofluorobenzene	11.085	95	126358	44.210	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.420%
Target Compounds						
11) Tert butyl alcohol	2.971	59	18644	28.539	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070622\
Data File : VX029942.D
Acq On : 06 Jul 2022 14:30
Operator : JC/MD
Sample : N3599-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EFF0722

Quant Time: Jul 07 02:36:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

