

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070622\
 Data File : VX029932.D
 Acq On : 06 Jul 2022 09:02
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 07 02:30:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	108	0.00
2 T	Dichlorodifluoromethane	50.000	49.292	1.4	97	0.00
3 P	Chloromethane	50.000	43.743	12.5	87	0.00
4 C	Vinyl Chloride	50.000	49.301	1.4#	97	0.00
5 T	Bromomethane	50.000	37.587	24.8	80	0.00
6 T	Chloroethane	50.000	41.579	16.8	90	0.00
7 T	Trichlorofluoromethane	50.000	49.545	0.9	100	0.00
8 T	Diethyl Ether	50.000	46.545	6.9	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.190	3.6	101	0.00
10 T	Methyl Iodide	50.000	35.619	28.8#	72	0.00
11 T	Tert butyl alcohol	250.000	227.867	8.9	94	0.00
12 CM	1,1-Dichloroethene	50.000	47.260	5.5#	98	0.00
13 T	Acrolein	250.000	110.196	55.9#	46	0.00
14 T	Allyl chloride	50.000	45.902	8.2	94	0.00
15 T	Acrylonitrile	250.000	229.189	8.3	93	0.00
16 T	Acetone	250.000	220.859	11.7	92	0.00
17 T	Carbon Disulfide	50.000	44.379	11.2	90	0.00
18 T	Methyl Acetate	50.000	45.068	9.9	94	0.00
19 T	Methyl tert-butyl Ether	50.000	47.202	5.6	95	0.00
20 T	Methylene Chloride	50.000	43.616	12.8	96	0.00
21 T	trans-1,2-Dichloroethene	50.000	46.020	8.0	96	0.00
22 T	Diisopropyl ether	50.000	46.132	7.7	93	0.00
23 T	Vinyl Acetate	250.000	232.799	6.9	92	0.00
24 P	1,1-Dichloroethane	50.000	47.631	4.7	97	0.00
25 T	2-Butanone	250.000	224.508	10.2	91	0.00
26 T	2,2-Dichloropropane	50.000	51.517	-3.0	104	0.00
27 T	cis-1,2-Dichloroethene	50.000	47.830	4.3	98	0.00
28 T	Bromochloromethane	50.000	46.571	6.9	96	0.00
29 T	Tetrahydrofuran	250.000	228.078	8.8	91	0.00
30 C	Chloroform	50.000	48.272	3.5#	98	0.00
31 T	Cyclohexane	50.000	48.103	3.8	99	0.00
32 T	1,1,1-Trichloroethane	50.000	48.284	3.4	99	0.00
33 S	1,2-Dichloroethane-d4	50.000	46.914	6.2	95	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	104	0.00
35 S	Dibromofluoromethane	50.000	48.988	2.0	98	0.00
36 T	1,1-Dichloropropene	50.000	48.427	3.1	98	0.00
37 T	Ethyl Acetate	50.000	45.857	8.3	90	0.00
38 T	Carbon Tetrachloride	50.000	50.950	-1.9	101	0.00
39 T	Methylcyclohexane	50.000	52.355	-4.7	103	0.00
40 TM	Benzene	50.000	48.605	2.8	97	0.00
41 T	Methacrylonitrile	50.000	47.476	5.0	92	0.00
42 TM	1,2-Dichloroethane	50.000	48.244	3.5	93	0.00
43 T	Isopropyl Acetate	50.000	47.777	4.4	93	0.00
44 TM	Trichloroethene	50.000	47.256	5.5	102	0.00
45 C	1,2-Dichloropropane	50.000	48.195	3.6#	96	0.00
46 T	Dibromomethane	50.000	49.593	0.8	97	0.00
47 T	Bromodichloromethane	50.000	50.362	-0.7	97	0.00
48 T	Methyl methacrylate	50.000	48.241	3.5	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070622\
 Data File : VX029932.D
 Acq On : 06 Jul 2022 09:02
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 07 02:30:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	973.292	2.7	94	0.00
50 S	Toluene-d8	50.000	51.150	-2.3	100	0.00
51 T	4-Methyl-2-Pentanone	250.000	238.926	4.4	92	0.00
52 CM	Toluene	50.000	49.900	0.2#	97	0.00
53 T	t-1,3-Dichloropropene	50.000	51.567	-3.1	97	0.00
54 T	cis-1,3-Dichloropropene	50.000	51.358	-2.7	98	0.00
55 T	1,1,2-Trichloroethane	50.000	47.986	4.0	95	0.00
56 T	Ethyl methacrylate	50.000	49.463	1.1	94	0.00
57 T	1,3-Dichloropropane	50.000	48.207	3.6	96	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	223.579	10.6	88	0.00
59 T	2-Hexanone	250.000	238.567	4.6	92	0.00
60 T	Dibromochloromethane	50.000	51.268	-2.5	98	0.00
61 T	1,2-Dibromoethane	50.000	48.093	3.8	95	0.00
62 S	4-Bromofluorobenzene	50.000	52.525	-5.0	101	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	104	0.00
64 T	Tetrachloroethene	50.000	58.701	-17.4	124	0.00
65 PM	Chlorobenzene	50.000	48.001	4.0	98	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	51.225	-2.5	99	0.00
67 C	Ethyl Benzene	50.000	49.435	1.1#	99	0.00
68 T	m/p-Xylenes	100.000	99.897	0.1	100	0.00
69 T	o-Xylene	50.000	48.380	3.2	97	0.00
70 T	Styrene	50.000	49.725	0.5	99	0.00
71 P	Bromoform	50.000	53.528	-7.1	98	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	109	0.00
73 T	Isopropylbenzene	50.000	46.718	6.6	101	0.00
74 T	N-amyl acetate	50.000	44.870	10.3	89	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	44.175	11.7	94	0.00
76 T	1,2,3-Trichloropropane	50.000	46.916	6.2	97	0.00
77 T	Bromobenzene	50.000	46.233	7.5	98	0.00
78 T	n-propylbenzene	50.000	48.112	3.8	101	0.00
79 T	2-Chlorotoluene	50.000	46.784	6.4	99	0.00
80 T	1,3,5-Trimethylbenzene	50.000	47.711	4.6	102	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	51.764	-3.5	98	0.00
82 T	4-Chlorotoluene	50.000	47.091	5.8	102	0.00
83 T	tert-Butylbenzene	50.000	47.013	6.0	100	0.00
84 T	1,2,4-Trimethylbenzene	50.000	48.685	2.6	103	0.00
85 T	sec-Butylbenzene	50.000	50.888	-1.8	107	0.00
86 T	p-Isopropyltoluene	50.000	51.959	-3.9	108	0.00
87 T	1,3-Dichlorobenzene	50.000	48.083	3.8	105	0.00
88 T	1,4-Dichlorobenzene	50.000	46.359	7.3	102	0.00
89 T	n-Butylbenzene	50.000	53.587	-7.2	113	0.00
90 T	Hexachloroethane	50.000	52.515	-5.0	110	0.00
91 T	1,2-Dichlorobenzene	50.000	48.110	3.8	103	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	49.718	0.6	105	0.00
93 T	1,2,4-Trichlorobenzene	50.000	48.176	3.6	103	0.00
94 T	Hexachlorobutadiene	50.000	48.094	3.8	107	0.00
95 T	Naphthalene	50.000	47.777	4.4	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070622\
Data File : VX029932.D
Acq On : 06 Jul 2022 09:02
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jul 07 02:30:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	47.942	4.1	102	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6