

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070622\
 Data File : VX029948.D
 Acq On : 06 Jul 2022 17:34
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 07 02:38:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	97	0.00
2 T	Dichlorodifluoromethane	50.000	50.613	-1.2	89	0.00
3 P	Chloromethane	50.000	43.354	13.3	78	0.00
4 C	Vinyl Chloride	50.000	49.824	0.4#	89	0.00
5 T	Bromomethane	50.000	28.806	42.4#	56	0.00
6 T	Chloroethane	50.000	39.660	20.7	78	0.00
7 T	Trichlorofluoromethane	50.000	51.428	-2.9	94	0.00
8 T	Diethyl Ether	50.000	49.452	1.1	92	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	49.763	0.5	94	0.00
10 T	Methyl Iodide	50.000	22.702	54.6#	40	0.00
11 T	Tert butyl alcohol	250.000	254.218	-1.7	94	0.00
12 CM	1,1-Dichloroethene	50.000	50.326	-0.7#	94	0.00
13 T	Acrolein	250.000	192.935	22.8	71	0.00
14 T	Allyl chloride	50.000	48.351	3.3	89	0.00
15 T	Acrylonitrile	250.000	242.590	3.0	89	0.00
16 T	Acetone	250.000	237.601	5.0	89	0.00
17 T	Carbon Disulfide	50.000	44.786	10.4	82	0.00
18 T	Methyl Acetate	50.000	49.021	2.0	92	0.00
19 T	Methyl tert-butyl Ether	50.000	49.412	1.2	89	0.00
20 T	Methylene Chloride	50.000	45.497	9.0	90	0.00
21 T	trans-1,2-Dichloroethene	50.000	47.635	4.7	90	0.00
22 T	Diisopropyl ether	50.000	48.834	2.3	88	0.00
23 T	Vinyl Acetate	250.000	245.452	1.8	87	0.00
24 P	1,1-Dichloroethane	50.000	48.993	2.0	90	0.00
25 T	2-Butanone	250.000	239.273	4.3	88	0.00
26 T	2,2-Dichloropropane	50.000	49.937	0.1	91	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.069	-0.1	93	0.00
28 T	Bromochloromethane	50.000	46.530	6.9	86	0.00
29 T	Tetrahydrofuran	250.000	245.480	1.8	89	0.00
30 C	Chloroform	50.000	50.336	-0.7#	92	0.00
31 T	Cyclohexane	50.000	49.479	1.0	91	0.00
32 T	1,1,1-Trichloroethane	50.000	50.579	-1.2	94	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.382	1.2	90	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	95	0.00
35 S	Dibromofluoromethane	50.000	51.551	-3.1	95	0.00
36 T	1,1-Dichloropropene	50.000	49.558	0.9	92	0.00
37 T	Ethyl Acetate	50.000	48.589	2.8	88	0.00
38 T	Carbon Tetrachloride	50.000	52.296	-4.6	95	0.00
39 T	Methylcyclohexane	50.000	52.423	-4.8	94	0.00
40 TM	Benzene	50.000	49.687	0.6	91	0.00
41 T	Methacrylonitrile	50.000	50.131	-0.3	89	0.00
42 TM	1,2-Dichloroethane	50.000	50.438	-0.9	89	0.00
43 T	Isopropyl Acetate	50.000	49.359	1.3	88	0.00
44 TM	Trichloroethene	50.000	48.908	2.2	96	0.00
45 C	1,2-Dichloropropane	50.000	49.867	0.3#	91	0.00
46 T	Dibromomethane	50.000	50.830	-1.7	91	0.00
47 T	Bromodichloromethane	50.000	51.424	-2.8	91	0.00
48 T	Methyl methacrylate	50.000	50.588	-1.2	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070622\
 Data File : VX029948.D
 Acq On : 06 Jul 2022 17:34
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 07 02:38:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1035.497	-3.5	92	0.00
50 S	Toluene-d8	50.000	52.407	-4.8	94	0.00
51 T	4-Methyl-2-Pentanone	250.000	250.853	-0.3	89	0.00
52 CM	Toluene	50.000	51.313	-2.6#	92	0.00
53 T	t-1,3-Dichloropropene	50.000	52.497	-5.0	90	0.00
54 T	cis-1,3-Dichloropropene	50.000	52.033	-4.1	91	0.00
55 T	1,1,2-Trichloroethane	50.000	50.638	-1.3	92	0.00
56 T	Ethyl methacrylate	50.000	50.696	-1.4	89	0.00
57 T	1,3-Dichloropropane	50.000	49.458	1.1	90	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	227.115	9.2	82	0.00
59 T	2-Hexanone	250.000	254.212	-1.7	90	0.00
60 T	Dibromochloromethane	50.000	52.901	-5.8	93	0.00
61 T	1,2-Dibromoethane	50.000	50.416	-0.8	92	0.00
62 S	4-Bromofluorobenzene	50.000	54.515	-9.0	96	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	95	0.00
64 T	Tetrachloroethene	50.000	56.868	-13.7	110	0.00
65 PM	Chlorobenzene	50.000	49.438	1.1	93	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	52.043	-4.1	92	0.00
67 C	Ethyl Benzene	50.000	50.554	-1.1#	93	0.00
68 T	m/p-Xylenes	100.000	102.587	-2.6	95	0.00
69 T	o-Xylene	50.000	50.404	-0.8	93	0.00
70 T	Styrene	50.000	51.522	-3.0	94	0.00
71 P	Bromoform	50.000	54.473	-8.9	92	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	99	0.00
73 T	Isopropylbenzene	50.000	48.890	2.2	96	0.00
74 T	N-amyl acetate	50.000	47.947	4.1	87	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	45.122	9.8	87	0.00
76 T	1,2,3-Trichloropropane	50.000	48.790	2.4	92	0.00
77 T	Bromobenzene	50.000	47.963	4.1	93	0.00
78 T	n-propylbenzene	50.000	50.444	-0.9	96	0.00
79 T	2-Chlorotoluene	50.000	48.613	2.8	94	0.00
80 T	1,3,5-Trimethylbenzene	50.000	50.578	-1.2	98	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	52.520	-5.0	90	0.00
82 T	4-Chlorotoluene	50.000	48.992	2.0	96	0.00
83 T	tert-Butylbenzene	50.000	49.666	0.7	96	0.00
84 T	1,2,4-Trimethylbenzene	50.000	49.305	1.4	95	0.00
85 T	sec-Butylbenzene	50.000	53.415	-6.8	102	0.00
86 T	p-Isopropyltoluene	50.000	53.865	-7.7	102	0.00
87 T	1,3-Dichlorobenzene	50.000	48.868	2.3	97	0.00
88 T	1,4-Dichlorobenzene	50.000	48.708	2.6	97	0.00
89 T	n-Butylbenzene	50.000	55.571	-11.1	106	0.00
90 T	Hexachloroethane	50.000	55.331	-10.7	105	0.00
91 T	1,2-Dichlorobenzene	50.000	49.377	1.2	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	51.038	-2.1	97	0.00
93 T	1,2,4-Trichlorobenzene	50.000	50.084	-0.2	98	0.00
94 T	Hexachlorobutadiene	50.000	50.990	-2.0	103	0.00
95 T	Naphthalene	50.000	50.183	-0.4	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070622\
Data File : VX029948.D
Acq On : 06 Jul 2022 17:34
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jul 07 02:38:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	49.613	0.8	96	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6