

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070623\
 Data File : VX036424.D
 Acq On : 06 Jul 2023 10:40
 Operator : JC/MD
 Sample : VX0706WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0706WBL01

Quant Time: Jul 07 01:43:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	169365	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	288444	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	262387	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	118340	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	108081	44.519	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.040%
35) Dibromofluoromethane	5.385	113	89138	46.867	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.740%
50) Toluene-d8	8.647	98	349255	49.166	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.340%
62) 4-Bromofluorobenzene	11.079	95	123347	49.296	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070623\
Data File : VX036424.D
Acq On : 06 Jul 2023 10:40
Operator : JC/MD
Sample : VX0706WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0706WBL01

Quant Time: Jul 07 01:43:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

