

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX070623\  
 Data File : VX036427.D  
 Acq On : 06 Jul 2023 11:55  
 Operator : JC/MD  
 Sample : 03500-08  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 7 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 FB-060923

Quant Time: Jul 07 01:45:16 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X062323W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Jun 23 15:03:28 2023  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon           | Response | Conc   | Units   | Dev(Min)   |
|-----------------------------|--------|----------------|----------|--------|---------|------------|
| -----                       |        |                |          |        |         |            |
| Internal Standards          |        |                |          |        |         |            |
| 1) Pentafluorobenzene       | 5.550  | 168            | 166879   | 50.000 | ug/l    | 0.00       |
| 34) 1,4-Difluorobenzene     | 6.763  | 114            | 278073   | 50.000 | ug/l    | 0.00       |
| 63) Chlorobenzene-d5        | 10.055 | 117            | 254650   | 50.000 | ug/l    | 0.00       |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152            | 113672   | 50.000 | ug/l    | 0.00       |
| System Monitoring Compounds |        |                |          |        |         |            |
| 33) 1,2-Dichloroethane-d4   | 5.958  | 65             | 105764   | 44.214 | ug/l    | 0.00       |
| Spiked Amount               | 50.000 | Range 78 - 117 | Recovery | =      | 88.420% |            |
| 35) Dibromofluoromethane    | 5.391  | 113            | 89051    | 48.568 | ug/l    | 0.00       |
| Spiked Amount               | 50.000 | Range 75 - 124 | Recovery | =      | 97.140% |            |
| 50) Toluene-d8              | 8.647  | 98             | 336002   | 49.064 | ug/l    | 0.00       |
| Spiked Amount               | 50.000 | Range 92 - 112 | Recovery | =      | 98.120% |            |
| 62) 4-Bromofluorobenzene    | 11.079 | 95             | 120457   | 49.937 | ug/l    | 0.00       |
| Spiked Amount               | 50.000 | Range 83 - 123 | Recovery | =      | 99.880% |            |
| Target Compounds            |        |                |          |        |         |            |
| 16) Acetone                 | 2.392  | 43             | 9371     | 9.165  | ug/l    | Qvalue 100 |
| -----                       |        |                |          |        |         |            |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX070623\  
Data File : VX036427.D  
Acq On : 06 Jul 2023 11:55  
Operator : JC/MD  
Sample : 03500-08  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 7 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
FB-060923

Quant Time: Jul 07 01:45:16 2023  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X062323W.M  
Quant Title : SW846 8260  
QLast Update : Fri Jun 23 15:03:28 2023  
Response via : Initial Calibration

