

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070623\
 Data File : VX036432.D
 Acq On : 06 Jul 2023 13:54
 Operator : JC/MD
 Sample : 03509-20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-WWS-04-G

Quant Time: Jul 07 01:47:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	165432	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	284861	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	258114	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	114281	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	107720	45.426	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.860%
35) Dibromofluoromethane	5.391	113	88839	47.298	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.600%
50) Toluene-d8	8.653	98	348351	49.655	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.320%
62) 4-Bromofluorobenzene	11.079	95	117950	47.732	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.460%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	225994	222.965	ug/l	94
18) Methyl Acetate	2.709	43	4194	1.006	ug/l	96
20) Methylene Chloride	2.788	84	13906	6.372	ug/l	94
25) 2-Butanone	4.568	43	116659	80.106	ug/l	94
37) Ethyl Acetate	4.721	43	38935	13.644	ug/l #	91
43) Isopropyl Acetate	6.342	43	142649	31.336	ug/l	96
51) 4-Methyl-2-Pentanone	8.574	43	456987	155.814	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070623\
Data File : VX036432.D
Acq On : 06 Jul 2023 13:54
Operator : JC/MD
Sample : 03509-20
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-WWS-04-G

Quant Time: Jul 07 01:47:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

