

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070623\
 Data File : VX036435.D
 Acq On : 06 Jul 2023 15:05
 Operator : JC/MD
 Sample : 03510-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-WWS-07-G

Quant Time: Jul 07 01:48:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	164061	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	277210	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	247628	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	107851	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	106549	45.307	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.620%
35) Dibromofluoromethane	5.385	113	85631	46.848	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.700%
50) Toluene-d8	8.653	98	331552	48.565	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.120%
62) 4-Bromofluorobenzene	11.079	95	115626	48.083	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.160%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.995	59	4784	11.176	ug/l	100
16) Acetone	2.392	43	239633	238.397	ug/l	93
18) Methyl Acetate	2.709	43	4875	1.179	ug/l	97
20) Methylene Chloride	2.788	84	8001	3.697	ug/l	92
25) 2-Butanone	4.562	43	119611	82.819	ug/l	99
37) Ethyl Acetate	4.727	43	6804	2.450	ug/l #	93
43) Isopropyl Acetate	6.342	43	32743	7.391	ug/l #	96
51) 4-Methyl-2-Pentanone	8.573	43	423950	148.539	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070623\
Data File : VX036435.D
Acq On : 06 Jul 2023 15:05
Operator : JC/MD
Sample : 03510-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-WWS-07-G

Quant Time: Jul 07 01:48:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

