

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070623\
 Data File : VX036439.D
 Acq On : 06 Jul 2023 16:40
 Operator : JC/MD
 Sample : 03514-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CP-5

Quant Time: Jul 07 01:50:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	169968	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	291132	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	261832	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	119482	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	112449	46.154	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	92.300%	
35) Dibromofluoromethane	5.391	113	90802	47.301	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	94.600%	
50) Toluene-d8	8.653	98	353648	49.324	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	98.640%	
62) 4-Bromofluorobenzene	11.079	95	124873	49.446	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	98.900%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.971	59	6486	14.625	ug/l #	95
16) Acetone	2.386	43	227025	218.005	ug/l	95
20) Methylene Chloride	2.788	84	12944	5.773	ug/l	96
25) 2-Butanone	4.562	43	124414	83.151	ug/l	98
37) Ethyl Acetate	4.721	43	52407	17.969	ug/l	97
43) Isopropyl Acetate	6.342	43	190067	40.852	ug/l	98
51) 4-Methyl-2-Pentanone	8.574	43	519235	173.225	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070623\
Data File : VX036439.D
Acq On : 06 Jul 2023 16:40
Operator : JC/MD
Sample : 03514-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CP-5

Quant Time: Jul 07 01:50:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

