

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070623\
 Data File : VX036440.D
 Acq On : 06 Jul 2023 17:28
 Operator : JC/MD
 Sample : 03500-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-3

Quant Time: Jul 07 01:50:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.555	168	170105	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	285743	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	255852	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	111095	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	109315	44.832	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.660%
35) Dibromofluoromethane	5.391	113	89242	47.366	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.740%
50) Toluene-d8	8.652	98	342688	48.697	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.400%
62) 4-Bromofluorobenzene	11.079	95	119254	48.111	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.220%
Target Compounds						
16) Acetone	2.385	43	2567	2.463	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070623\
Data File : VX036440.D
Acq On : 06 Jul 2023 17:28
Operator : JC/MD
Sample : 03500-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-3

Quant Time: Jul 07 01:50:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

