

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070720\
 Data File : VX017195.D
 Acq On : 07 Jul 2020 12:04
 Operator : JC/SP
 Sample : L3210-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT-TB-20200701

Quant Time: Jul 07 14:54:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	254114	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	426319	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	430680	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	228956	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	165910	52.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.92%	
35) Dibromofluoromethane	5.46	113	141245	52.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.76%	
50) Toluene-d8	8.70	98	515497	51.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.38%	
62) 4-Bromofluorobenzene	11.12	95	207196	53.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.30%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070720\
Data File : VX017195.D
Acq On : 07 Jul 2020 12:04
Operator : JC/SP
Sample : L3210-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-TB-20200701

Quant Time: Jul 07 14:54:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

