

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070720\
 Data File : VX017197.D
 Acq On : 07 Jul 2020 12:50
 Operator : JC/SP
 Sample : L3159-25
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB06-20200629

Quant Time: Jul 08 05:14:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	257092	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	428390	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	428936	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	228317	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	166031	52.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.78%	
35) Dibromofluoromethane	5.47	113	142142	52.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.90%	
50) Toluene-d8	8.70	98	512957	51.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.38%	
62) 4-Bromofluorobenzene	11.12	95	208413	53.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.42%	
Target Compounds						
16) Acetone	2.42	43	9985	8.422	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070720\
Data File : VX017197.D
Acq On : 07 Jul 2020 12:50
Operator : JC/SP
Sample : L3159-25
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB06-20200629

Quant Time: Jul 08 05:14:55 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

