

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070720\
 Data File : VX017200.D
 Acq On : 07 Jul 2020 13:58
 Operator : JC/SP
 Sample : L3159-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 23BR-20200629

Quant Time: Jul 08 05:48:00 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	258623	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	426636	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	434741	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	237011	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	167178	52.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.86%	
35) Dibromofluoromethane	5.47	113	142351	52.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.50%	
50) Toluene-d8	8.70	98	517260	51.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.66%	
62) 4-Bromofluorobenzene	11.12	95	213887	54.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.66%	
Target Compounds						
16) Acetone	2.42	43	7333	6.148	ug/l #	87
27) cis-1,2-Dichloroethene	4.57	96	27945	8.686	ug/l	90
44) Trichloroethene	7.19	130	109889	33.511	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070720\
Data File : VX017200.D
Acq On : 07 Jul 2020 13:58
Operator : JC/SP
Sample : L3159-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
23BR-20200629

Quant Time: Jul 08 05:48:00 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

