

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX070720\
 Data File : VX017213.D
 Acq On : 07 Jul 2020 18:54
 Operator : JC/SP
 Sample : L3234-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-3

Manual Integrations
 APPROVED

MMDadoda
 7/8/2020 9:31:39 AM

Quant Time: Jul 10 16:52:48 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	258609	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	424717	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	432963	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	241557	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	167166	52.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.86%	
35) Dibromofluoromethane	5.47	113	144333	53.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.44%	
50) Toluene-d8	8.70	98	518771	52.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.44%	
62) 4-Bromofluorobenzene	11.12	95	224964	57.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.86%	
Target Compounds						
					Qvalue	
16) Acetone	2.42	43	3554	2.980	ug/l	98
31) Cyclohexane	5.56	56	27377	6.360	ug/l	91
39) Methylcyclohexane	7.45	83	5133	1.149	ug/l	91
40) Benzene	6.12	78	14727	1.249	ug/l	97
52) Toluene	8.76	92	83149	11.292	ug/l	99
67) Ethyl Benzene	10.24	91	542987	36.385	ug/l	98
68) m/p-Xylenes	10.34	106	339141	59.186	ug/l	98
69) o-Xylene	10.68	106	131213	24.071	ug/l	98
73) Isopropylbenzene	11.00	105	75649	4.781	ug/l	97
78) n-propylbenzene	11.34	91	136714	7.641	ug/l	100
80) 1,3,5-Trimethylbenzene	11.49	105	33924	2.585	ug/l	99
84) 1,2,4-Trimethylbenzene	11.79	105	224111	16.832	ug/l	99
85) sec-Butylbenzene	11.93	105	4029m	0.266	ug/l	
95) Naphthalene	13.82	128	87015	5.680	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX070720\
Data File : VX017213.D
Acq On : 07 Jul 2020 18:54
Operator : JC/SP
Sample : L3234-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-3

Manual Integrations
APPROVED

MMDadoda
7/8/2020 9:31:39 AM

Quant Time: Jul 10 16:52:48 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

