

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070720\
 Data File : VX017225.D
 Acq On : 08 Jul 2020 00:35
 Operator : JC/SP
 Sample : L3192-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-MW180-SOIL-600-604

Quant Time: Jul 08 06:50:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	244561	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	417454	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	452925	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	232814	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	161066	53.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.84%	
35) Dibromofluoromethane	5.47	113	141212	53.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.96%	
50) Toluene-d8	8.70	98	522857	53.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.10%	
62) 4-Bromofluorobenzene	11.12	95	220689	57.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.64%	
Target Compounds						
16) Acetone	2.42	43	21592	19.144	ug/l	98
18) Methyl Acetate	2.75	43	3500	1.189	ug/l #	89
20) Methylene Chloride	2.84	84	20528	7.267	ug/l	95
43) Isopropyl Acetate	6.42	43	75955	11.418	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070720\
Data File : VX017225.D
Acq On : 08 Jul 2020 00:35
Operator : JC/SP
Sample : L3192-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-MW180-SOIL-600-604

Quant Time: Jul 08 06:50:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

