

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070720\
 Data File : VX017230.D
 Acq On : 08 Jul 2020 02:29
 Operator : JC/SP
 Sample : L3232-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1014-TB070620

Quant Time: Jul 08 04:20:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	245553	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	420318	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	434237	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	232593	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	165801	54.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.54%	
35) Dibromofluoromethane	5.47	113	147184	55.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.72%	
50) Toluene-d8	8.70	98	514589	52.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.68%	
62) 4-Bromofluorobenzene	11.12	95	208716	54.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.62%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070720\
Data File : VX017230.D
Acq On : 08 Jul 2020 02:29
Operator : JC/SP
Sample : L3232-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1014-TB070620

Quant Time: Jul 08 04:20:55 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

