

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070720\
 Data File : VX017236.D
 Acq On : 08 Jul 2020 04:46
 Operator : JC/SP
 Sample : L3232-11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1011-MW-18

Quant Time: Jul 08 07:08:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	244808	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	417009	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	436062	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	233058	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	165837	54.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.90%	
35) Dibromofluoromethane	5.46	113	142586	54.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.10%	
50) Toluene-d8	8.70	98	508359	52.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.24%	
62) 4-Bromofluorobenzene	11.12	95	210285	55.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.30%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.57	96	5056	1.660	ug/l	83
44) Trichloroethene	7.20	130	7202	1.956	ug/l	90
64) Tetrachloroethene	9.32	164	97786	27.786	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070720\
Data File : VX017236.D
Acq On : 08 Jul 2020 04:46
Operator : JC/SP
Sample : L3232-11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1011-MW-18

Quant Time: Jul 08 07:08:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

