

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070720\  
 Data File : VX017239.D  
 Acq On : 08 Jul 2020 05:54  
 Operator : JC/SP  
 Sample : L3232-07  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 53 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 1007-MW-11

Quant Time: Jul 08 07:12:13 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X062920W.M  
 Quant Title : SW846 8260  
 QLast Update : Mon Jun 29 17:10:38 2020  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc    | Units   | Dev(Min) |
|-----------------------------|-------|------|----------|---------|---------|----------|
| 1) Pentafluorobenzene       | 5.63  | 168  | 253652   | 50.00   | ug/l    | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.83  | 114  | 429943   | 50.00   | ug/l    | 0.00     |
| 63) Chlorobenzene-d5        | 10.10 | 117  | 453724   | 50.00   | ug/l    | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.06 | 152  | 239306   | 50.00   | ug/l    | 0.00     |
| System Monitoring Compounds |       |      |          |         |         |          |
| 33) 1,2-Dichloroethane-d4   | 6.04  | 65   | 170176   | 54.42   | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =       | 108.84% |          |
| 35) Dibromofluoromethane    | 5.47  | 113  | 148974   | 54.78   | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =       | 109.56% |          |
| 50) Toluene-d8              | 8.70  | 98   | 531621   | 52.86   | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =       | 105.72% |          |
| 62) 4-Bromofluorobenzene    | 11.12 | 95   | 218498   | 55.58   | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =       | 111.16% |          |
| Target Compounds            |       |      |          |         |         |          |
| 16) Acetone                 | 2.42  | 43   | 12127    | 10.367  | ug/l    | 100      |
| 25) 2-Butanone              | 4.66  | 43   | 4449     | 2.283   | ug/l #  | 81       |
| 27) cis-1,2-Dichloroethene  | 4.56  | 96   | 25440    | 8.062   | ug/l    | 90       |
| 44) Trichloroethene         | 7.19  | 130  | 42530    | 12.678  | ug/l    | 98       |
| 64) Tetrachloroethene       | 9.32  | 164  | 715311   | 195.342 | ug/l    | 97       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070720\  
Data File : VX017239.D  
Acq On : 08 Jul 2020 05:54  
Operator : JC/SP  
Sample : L3232-07  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 53 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
1007-MW-11

Quant Time: Jul 08 07:12:13 2020  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_X\METHOD\82X062920W.M  
Quant Title : SW846 8260  
QLast Update : Mon Jun 29 17:10:38 2020  
Response via : Initial Calibration

