

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070720\
 Data File : VX017195.D
 Acq On : 07 Jul 2020 12:04
 Operator : JC/SP
 Sample : L3210-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT-TB-20200701

Quant Time: Jul 07 14:54:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	254114	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	426319	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	430680	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	228956	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	165910	52.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.92%	
35) Dibromofluoromethane	5.46	113	141245	52.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.76%	
50) Toluene-d8	8.70	98	515497	51.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.38%	
62) 4-Bromofluorobenzene	11.12	95	207196	53.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.30%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070720\
Data File : VX017195.D
Acq On : 07 Jul 2020 12:04
Operator : JC/SP
Sample : L3210-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-TB-20200701

Quant Time: Jul 07 14:54:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.63 min Scan# 745

Delta R.T. -0.01 min

Lab File: VX017195.D

Acq: 07 Jul 2020 12:04

Instrument :

MSVOA_X

ClientSampleId :


TT-TB-20200701

Tot Ion:168 Resp: 254114

Ion Ratio Lower Upper

168 100

99 54.0 42.6 64.0



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.63

100000

80000

60000



40000

20000

0

Time-->

5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 52.960 ug/l

RT: 6.03 min Scan# 811

Delta R.T. -0.00 min

Lab File: VX017195.D


Acq: 07 Jul 2020 12:04

Tgt Ion: 65 Resp: 165910

Ion Ratio Lower Upper

65 100

67 51.6 0.0 103.8



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.03

60000


40000


20000

0

Time-->

5.90 6.00 6.10





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.83 min Scan# 942

Delta R.T. -0.00 min

Lab File: VX017195.D

Acq: 07 Jul 2020 12:04

Instrument :

MSVOA_X

ClientSampleId :

TT-TB-20200701


Tot Ion:114 Resp: 426319

Ion Ratio Lower Upper

114 100

63 20.5 0.0 41.4




88 15.9 0.0 32.4



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 52.376 ug/l

RT: 5.46 min Scan# 718

Delta R.T. -0.00 min

Lab File: VX017195.D

Acq: 07 Jul 2020 12:04


Tgt Ion:113 Resp: 141245

Ion Ratio Lower Upper

113 100

111 102.6 81.9 122.9



192 21.9 16.7 25.1




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 51.694 ug/l

RT: 8.70 min Scan# 1248

Delta R.T. -0.00 min

Lab File: VX017195.D

Acq: 07 Jul 2020 12:04

Instrument :

MSVOA_X

ClientSampleId :





TT-TB-20200701

Tot Ion: 98 Resp: 515497

Ion Ratio Lower Upper

98 100

100 66.4 53.8 80.6



#62

4-Bromofluorobenzene

Concen: 53.152 ug/l

RT: 11.12 min Scan# 1646

Delta R.T. -0.00 min

Lab File: VX017195.D

Acq: 07 Jul 2020 12:04




Tgt Ion: 95 Resp: 207196


Ion Ratio Lower Upper

95 100

174 85.6 0.0 164.0

176 81.6 0.0 157.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.10 min Scan# 1478
Delta R.T. -0.00 min
Lab File: VX017195.D
Acq: 07 Jul 2020 12:04

Instrument :
MSVOA_X
ClientSampleId :
TT-TB-20200701

Tot Ion	Ratio	Lower	Upper
117	100		
82	51.9	42.6	63.8
119	31.7	25.4	38.2



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.06 min Scan# 1800
Delta R.T. -0.00 min
Lab File: VX017195.D
Acq: 07 Jul 2020 12:04

Tgt Ion	Ratio	Lower	Upper
152	100		
115	57.0	40.9	122.7
150	156.0	0.0	351.6

