

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070721\
 Data File : VX023160.D
 Acq On : 07 Jul 2021 10:12
 Operator : JC/MD
 Sample : VX0707WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0707WBL01

Quant Time: Jul 08 03:15:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	230612	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	383625	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	340568	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	148395	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	160106	51.657	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.320%
35) Dibromofluoromethane	5.397	113	122780	49.562	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.120%
50) Toluene-d8	8.653	98	453735	49.101	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.200%
62) 4-Bromofluorobenzene	11.085	95	165970	46.523	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.040%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070721\
Data File : VX023160.D
Acq On : 07 Jul 2021 10:12
Operator : JC/MD
Sample : VX0707WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0707WBL01

Quant Time: Jul 08 03:15:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

