

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070721\
 Data File : VX023164.D
 Acq On : 07 Jul 2021 11:56
 Operator : JC/MD
 Sample : M2923-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-12A-070121-FD

Quant Time: Jul 08 03:17:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	229848	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	380502	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	340469	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	144564	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	159021	51.477	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	102.960%	
35) Dibromofluoromethane	5.397	113	122211	49.737	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	99.480%	
50) Toluene-d8	8.653	98	451188	49.226	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	98.460%	
62) 4-Bromofluorobenzene	11.085	95	165839	46.868	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	93.740%	
Target Compounds						
						Qvalue
16) Acetone	2.392	43	3207	2.624	ug/l #	82
27) cis-1,2-Dichloroethene	4.501	96	8397	3.133	ug/l	92
44) Trichloroethene	7.135	130	48679	17.424	ug/l	98
64) Tetrachloroethene	9.281	164	888	0.327	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070721\
Data File : VX023164.D
Acq On : 07 Jul 2021 11:56
Operator : JC/MD
Sample : M2923-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-12A-070121-FD

Quant Time: Jul 08 03:17:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

