

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070721\
 Data File : VX023165.D
 Acq On : 07 Jul 2021 12:19
 Operator : JC/MD
 Sample : M2939-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-070221

Quant Time: Jul 08 03:17:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	223700	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	368439	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	329385	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	137601	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	152982	50.883	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.760%
35) Dibromofluoromethane	5.397	113	114681	48.201	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.400%
50) Toluene-d8	8.653	98	439627	49.535	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.080%
62) 4-Bromofluorobenzene	11.085	95	156634	45.716	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.440%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	2977	2.503	ug/l #	87
20) Methylene Chloride	2.800	84	1394	0.544	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070721\
Data File : VX023165.D
Acq On : 07 Jul 2021 12:19
Operator : JC/MD
Sample : M2939-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-070221

Quant Time: Jul 08 03:17:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

