

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070721\
 Data File : VX023166.D
 Acq On : 07 Jul 2021 12:43
 Operator : JC/MD
 Sample : M2939-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-01-070221

Quant Time: Jul 08 03:18:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	230943	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	382774	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	340485	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	143876	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	160955	51.856	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.720%
35) Dibromofluoromethane	5.397	113	121280	49.066	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.140%
50) Toluene-d8	8.653	98	449589	48.760	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.520%
62) 4-Bromofluorobenzene	11.085	95	164873	46.318	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.640%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	19417	15.812	ug/l #	88
20) Methylene Chloride	2.794	84	8886	3.361	ug/l	98
25) 2-Butanone	4.574	43	5174	2.794	ug/l	91
29) Tetrahydrofuran	5.032	42	2934	2.465	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070721\
Data File : VX023166.D
Acq On : 07 Jul 2021 12:43
Operator : JC/MD
Sample : M2939-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-01-070221

Quant Time: Jul 08 03:18:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

