

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070721\
 Data File : VX023168.D
 Acq On : 07 Jul 2021 13:30
 Operator : JC/MD
 Sample : M2939-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-04-07-12-070221

Quant Time: Jul 08 03:18:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	225960	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	373131	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	334924	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	135164	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.971	65	156769	51.621	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.240%
35) Dibromofluoromethane	5.398	113	117354	48.704	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.400%
50) Toluene-d8	8.653	98	447460	49.784	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.560%
62) 4-Bromofluorobenzene	11.086	95	156702	45.160	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.320%
Target Compounds						
16) Acetone	2.392	43	3214	2.675	ug/l #	80
20) Methylene Chloride	2.794	84	1194	0.462	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070721\
Data File : VX023168.D
Acq On : 07 Jul 2021 13:30
Operator : JC/MD
Sample : M2939-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-04-07-12-070221

Quant Time: Jul 08 03:18:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

