

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070721\
 Data File : VX023169.D
 Acq On : 07 Jul 2021 13:54
 Operator : JC/MD
 Sample : M2939-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-04-12-17-070221

Quant Time: Jul 08 03:18:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	224773	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	366142	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	331405	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	140162	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	154890	51.272	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.540%
35) Dibromofluoromethane	5.397	113	116322	49.197	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.400%
50) Toluene-d8	8.653	98	439997	49.888	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.780%
62) 4-Bromofluorobenzene	11.085	95	161484	47.427	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.860%
Target Compounds						
16) Acetone	2.385	43	3828	3.203	ug/l #	87
20) Methylene Chloride	2.794	84	1020	0.396	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070721\
Data File : VX023169.D
Acq On : 07 Jul 2021 13:54
Operator : JC/MD
Sample : M2939-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-04-12-17-070221

Quant Time: Jul 08 03:18:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

