

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070721\
 Data File : VX023170.D
 Acq On : 07 Jul 2021 14:17
 Operator : JC/MD
 Sample : M2939-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-BR2-070221

Quant Time: Jul 08 03:19:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	230923	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	376587	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	334051	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	138470	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	161210	51.943	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 103.880%		
35) Dibromofluoromethane	5.397	113	120204	49.429	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 98.860%		
50) Toluene-d8	8.653	98	441526	48.673	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 97.340%		
62) 4-Bromofluorobenzene	11.085	95	160344	45.786	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 91.580%		
Target Compounds						
						Qvalue
4) Vinyl Chloride	1.374	62	1853	0.845	ug/l	94
16) Acetone	2.392	43	4883	3.977	ug/l	95
17) Carbon Disulfide	2.520	76	1812	0.417	ug/l	95
20) Methylene Chloride	2.794	84	1169	0.442	ug/l #	87
24) 1,1-Dichloroethane	3.617	63	2308	0.542	ug/l	94
27) cis-1,2-Dichloroethene	4.501	96	21699	8.058	ug/l	93
52) Toluene	8.726	92	25191	3.993	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070721\
Data File : VX023170.D
Acq On : 07 Jul 2021 14:17
Operator : JC/MD
Sample : M2939-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-BR2-070221

Quant Time: Jul 08 03:19:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

