

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070721\
 Data File : VX023171.D
 Acq On : 07 Jul 2021 14:41
 Operator : JC/MD
 Sample : M2939-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-14A-070221

Quant Time: Jul 08 03:19:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	218262	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	361001	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	315637	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	131031	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	152784	52.084	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.160%
35) Dibromofluoromethane	5.397	113	114323	49.041	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.080%
50) Toluene-d8	8.653	98	424078	48.768	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.540%
62) 4-Bromofluorobenzene	11.085	95	148595	44.263	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.520%
Target Compounds						
16) Acetone	2.392	43	3146	2.711	ug/l	Qvalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070721\
Data File : VX023171.D
Acq On : 07 Jul 2021 14:41
Operator : JC/MD
Sample : M2939-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-14A-070221

Quant Time: Jul 08 03:19:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

