

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070721\
 Data File : VX023172.D
 Acq On : 07 Jul 2021 15:04
 Operator : JC/MD
 Sample : M2939-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-14B-070221

Quant Time: Jul 08 03:20:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	210481	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	347281	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	294783	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	111433	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	145680	51.498	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.000%
35) Dibromofluoromethane	5.397	113	110794	49.404	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.800%
50) Toluene-d8	8.653	98	406013	48.535	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.060%
62) 4-Bromofluorobenzene	11.085	95	136291	42.202	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.400%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	2477	2.213	ug/l	89
30) Chloroform	5.111	83	1844	0.440	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070721\
Data File : VX023172.D
Acq On : 07 Jul 2021 15:04
Operator : JC/MD
Sample : M2939-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-14B-070221

Quant Time: Jul 08 03:20:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

