

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070721\
 Data File : VX023173.D
 Acq On : 07 Jul 2021 15:28
 Operator : JC/MD
 Sample : M2948-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-3

Quant Time: Jul 08 03:20:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	214547	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	355955	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	315349	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	134901	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	152532	52.898	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.800%
35) Dibromofluoromethane	5.397	113	113427	49.346	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.700%
50) Toluene-d8	8.653	98	419750	48.954	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.900%
62) 4-Bromofluorobenzene	11.085	95	152076	45.942	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.880%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	2742	2.404	ug/l #	79
19) Methyl tert-butyl Ether	3.123	73	7175	1.003	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070721\
Data File : VX023173.D
Acq On : 07 Jul 2021 15:28
Operator : JC/MD
Sample : M2948-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-3

Quant Time: Jul 08 03:20:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

