

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070721\
 Data File : VX023174.D
 Acq On : 07 Jul 2021 15:51
 Operator : JC/MD
 Sample : M2948-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1

Quant Time: Jul 08 03:20:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	219596	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	365632	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	320656	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	134009	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	153118	51.880	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.760%
35) Dibromofluoromethane	5.397	113	116556	49.365	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.740%
50) Toluene-d8	8.653	98	429249	48.737	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.480%
62) 4-Bromofluorobenzene	11.085	95	153354	45.102	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.200%
Target Compounds						
16) Acetone	2.392	43	2350	2.013	ug/l	Qvalue # 75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070721\
Data File : VX023174.D
Acq On : 07 Jul 2021 15:51
Operator : JC/MD
Sample : M2948-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-1

Quant Time: Jul 08 03:20:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

