

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070721\
 Data File : VX023177.D
 Acq On : 07 Jul 2021 17:02
 Operator : JC/MD
 Sample : M2953-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-01-070621

Quant Time: Jul 08 03:21:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	212718	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	349166	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	297190	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	106659	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	147328	51.533	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 103.060%		
35) Dibromofluoromethane	5.397	113	110851	49.163	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 98.320%		
50) Toluene-d8	8.653	98	412497	49.044	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 98.080%		
62) 4-Bromofluorobenzene	11.085	95	135909	41.856	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 83.720%		
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	18391	16.260	ug/l	89
20) Methylene Chloride	2.794	84	9313	3.824	ug/l	92
25) 2-Butanone	4.574	43	5126	3.005	ug/l	89
27) cis-1,2-Dichloroethene	4.501	96	2747	1.107	ug/l	91
29) Tetrahydrofuran	5.025	42	2687	2.451	ug/l	88
44) Trichloroethene	7.135	130	28264	11.025	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070721\
Data File : VX023177.D
Acq On : 07 Jul 2021 17:02
Operator : JC/MD
Sample : M2953-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-01-070621

Quant Time: Jul 08 03:21:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

