

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070721\
 Data File : VX023181.D
 Acq On : 07 Jul 2021 18:37
 Operator : JC/MD
 Sample : M2953-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-BR3-070621

Quant Time: Jul 08 03:23:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	222043	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	365486	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	324921	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	135580	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	156169	52.331	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	104.660%	
35) Dibromofluoromethane	5.397	113	116861	49.514	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	99.020%	
50) Toluene-d8	8.653	98	431429	49.004	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	98.000%	
62) 4-Bromofluorobenzene	11.085	95	155468	45.742	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	91.480%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.380	62	837	0.397	ug/l #	79
9) 1,1,2-Trichlorotrifluo...	2.331	101	1380	0.643	ug/l	91
12) 1,1-Dichloroethene	2.325	96	68403	34.559	ug/l	93
16) Acetone	2.386	43	5507	4.664	ug/l #	84
17) Carbon Disulfide	2.514	76	22580	5.409	ug/l	99
20) Methylene Chloride	2.800	84	1734	0.682	ug/l	86
21) trans-1,2-Dichloroethene	3.099	96	4717	2.207	ug/l	97
24) 1,1-Dichloroethane	3.617	63	19209	4.688	ug/l	98
27) cis-1,2-Dichloroethene	4.495	96	67969	26.251	ug/l	93
30) Chloroform	5.111	83	3462	0.782	ug/l	99
42) 1,2-Dichloroethane	6.099	62	13495	3.550	ug/l	95
44) Trichloroethene	7.135	130	401143	149.484	ug/l	98
49) 1,4-Dioxane	7.671	88	1396	21.382	ug/l #	90
52) Toluene	8.720	92	4701	0.768	ug/l	100
64) Tetrachloroethene	9.275	164	1922	0.742	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070721\
Data File : VX023181.D
Acq On : 07 Jul 2021 18:37
Operator : JC/MD
Sample : M2953-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-BR3-070621

Quant Time: Jul 08 03:23:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

