

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070721\
 Data File : VX023158.D
 Acq On : 07 Jul 2021 09:15
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 08 03:13:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	104	0.00
2 T	Dichlorodifluoromethane	0.453	0.491	-8.4	105	0.00
3 P	Chloromethane	0.448	0.434	3.1	101	0.00
4 C	Vinyl Chloride	0.475	0.470	1.1#	100	0.00
5 T	Bromomethane	0.265	0.208	21.5	87	0.00
6 T	Chloroethane	0.310	0.289	6.8	94	0.00
7 T	Trichlorofluoromethane	0.825	0.784	5.0	95	0.00
8 T	Diethyl Ether	0.305	0.287	5.9	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.483	0.491	-1.7	104	0.00
10 T	Methyl Iodide	0.593	0.550	7.3	95	0.00
11 T	Tert butyl alcohol	0.141	0.128	9.2	99	0.00
12 CM	1,1-Dichloroethene	0.446	0.442	0.9#	100	0.00
13 T	Acrolein	0.094	0.075	20.2	83	0.00
14 T	Allyl chloride	0.774	0.794	-2.6	104	0.00
15 T	Acrylonitrile	0.275	0.264	4.0	97	0.00
16 T	Acetone	0.266	0.313	-17.7	128	0.00
17 T	Carbon Disulfide	0.940	0.960	-2.1	107	0.00
18 T	Methyl Acetate	0.645	0.612	5.1	97	0.00
19 T	Methyl tert-butyl Ether	1.667	1.679	-0.7	102	0.00
20 T	Methylene Chloride	0.572	0.505	11.7	99	0.00
21 T	trans-1,2-Dichloroethene	0.481	0.469	2.5	101	0.00
22 T	Diisopropyl ether	1.633	1.608	1.5	100	0.00
23 T	Vinyl Acetate	1.382	1.394	-0.9	99	0.00
24 P	1,1-Dichloroethane	0.923	0.918	0.5	102	0.00
25 T	2-Butanone	0.401	0.425	-6.0	108	0.00
26 T	2,2-Dichloropropane	0.774	0.859	-11.0	112	0.00
27 T	cis-1,2-Dichloroethene	0.583	0.570	2.2	101	0.00
28 T	Bromochloromethane	0.438	0.402	8.2	97	0.00
29 T	Tetrahydrofuran	0.258	0.243	5.8	96	0.00
30 C	Chloroform	0.997	0.980	1.7#	101	0.00
31 T	Cyclohexane	0.805	0.773	4.0	101	0.00
32 T	1,1,1-Trichloroethane	0.868	0.893	-2.9	104	0.00
33 S	1,2-Dichloroethane-d4	0.672	0.635	5.5	102	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	102	0.00
35 S	Dibromofluoromethane	0.323	0.320	0.9	106	0.00
36 T	1,1-Dichloropropene	0.446	0.449	-0.7	102	0.00
37 T	Ethyl Acetate	0.486	0.482	0.8	98	0.00
38 T	Carbon Tetrachloride	0.465	0.510	-9.7	106	0.00
39 T	Methylcyclohexane	0.521	0.542	-4.0	105	0.00
40 TM	Benzene	1.294	1.293	0.1	99	0.00
41 T	Methacrylonitrile	0.272	0.267	1.8	98	0.00
42 TM	1,2-Dichloroethane	0.520	0.527	-1.3	99	0.00
43 T	Isopropyl Acetate	0.809	0.801	1.0	98	0.00
44 TM	Trichloroethene	0.367	0.369	-0.5	100	0.00
45 C	1,2-Dichloropropane	0.338	0.344	-1.8#	102	0.00
46 T	Dibromomethane	0.241	0.242	-0.4	99	0.00
47 T	Bromodichloromethane	0.441	0.477	-8.2	104	0.00
48 T	Methyl methacrylate	0.397	0.392	1.3	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070721\
 Data File : VX023158.D
 Acq On : 07 Jul 2021 09:15
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 08 03:13:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	95	0.00
50 S	Toluene-d8	1.204	1.175	2.4	102	0.00
51 T	4-Methyl-2-Pentanone	0.501	0.496	1.0	96	0.00
52 CM	Toluene	0.838	0.852	-1.7#	100	0.00
53 T	t-1,3-Dichloropropene	0.481	0.542	-12.7	105	0.00
54 T	cis-1,3-Dichloropropene	0.510	0.571	-12.0	104	0.00
55 T	1,1,2-Trichloroethane	0.352	0.358	-1.7	100	0.00
56 T	Ethyl methacrylate	0.543	0.564	-3.9	101	0.00
57 T	1,3-Dichloropropane	0.581	0.576	0.9	98	0.00
58 T	2-Chloroethyl Vinyl ether	0.266	0.285	-7.1	98	0.00
59 T	2-Hexanone	0.385	0.403	-4.7	101	0.00
60 T	Dibromochloromethane	0.338	0.375	-10.9	105	0.00
61 T	1,2-Dibromoethane	0.361	0.370	-2.5	97	0.00
62 S	4-Bromofluorobenzene	0.465	0.465	0.0	103	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	99	0.00
64 T	Tetrachloroethene	0.399	0.403	-1.0	99	0.00
65 PM	Chlorobenzene	0.988	1.013	-2.5	100	0.00
66 T	1,1,1,2-Tetrachloroethane	0.357	0.388	-8.7	101	0.00
67 C	Ethyl Benzene	1.747	1.831	-4.8#	100	0.00
68 T	m/p-Xylenes	0.667	0.712	-6.7	101	0.00
69 T	o-Xylene	0.672	0.701	-4.3	100	0.00
70 T	Styrene	1.098	1.193	-8.7	101	0.00
71 P	Bromoform	0.252	0.293	-16.3	106	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
73 T	Isopropylbenzene	3.522	3.710	-5.3	100	0.00
74 T	N-amyl acetate	1.495	1.527	-2.1	97	0.00
75 P	1,1,2,2-Tetrachloroethane	1.119	1.112	0.6	98	0.00
76 T	1,2,3-Trichloropropane	1.006	1.068	-6.2	100	0.00
77 T	Bromobenzene	0.879	0.894	-1.7	99	0.00
78 T	n-propylbenzene	4.073	4.317	-6.0	100	0.00
79 T	2-Chlorotoluene	2.517	2.584	-2.7	99	0.00
80 T	1,3,5-Trimethylbenzene	3.033	3.212	-5.9	100	0.00
81 T	trans-1,4-Dichloro-2-butene	0.328	0.384	-17.1	110	0.00
82 T	4-Chlorotoluene	2.843	3.029	-6.5	100	0.00
83 T	tert-Butylbenzene	2.879	3.066	-6.5	100	0.00
84 T	1,2,4-Trimethylbenzene	2.997	3.208	-7.0	100	0.00
85 T	sec-Butylbenzene	3.704	3.999	-8.0	101	0.00
86 T	p-Isopropyltoluene	3.107	3.388	-9.0	102	0.00
87 T	1,3-Dichlorobenzene	1.651	1.702	-3.1	99	0.00
88 T	1,4-Dichlorobenzene	1.663	1.699	-2.2	99	0.00
89 T	n-Butylbenzene	2.722	2.970	-9.1	101	0.00
90 T	Hexachloroethane	0.455	0.548	-20.4	109	0.00
91 T	1,2-Dichlorobenzene	1.643	1.677	-2.1	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.238	0.270	-13.4	102	0.00
93 T	1,2,4-Trichlorobenzene	1.012	1.081	-6.8	102	0.00
94 T	Hexachlorobutadiene	0.428	0.480	-12.1	109	0.00
95 T	Naphthalene	3.366	3.643	-8.2	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070721\
Data File : VX023158.D
Acq On : 07 Jul 2021 09:15
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jul 08 03:13:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.018	1.093	-7.4	100	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6