

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070723\
 Data File : VX036453.D
 Acq On : 07 Jul 2023 12:18
 Operator : JC/MD
 Sample : 03533-01 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 7623

Quant Time: Jul 07 23:03:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	170129	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	288583	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	259985	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	115122	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	111883	45.878	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.760%
35) Dibromofluoromethane	5.385	113	90843	47.741	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.480%
50) Toluene-d8	8.647	98	348311	49.009	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.020%
62) 4-Bromofluorobenzene	11.079	95	122643	48.991	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.980%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	3101	2.975	ug/l	100
20) Methylene Chloride	2.788	84	1496	0.667	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070723\
Data File : VX036453.D
Acq On : 07 Jul 2023 12:18
Operator : JC/MD
Sample : 03533-01 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
7623

Quant Time: Jul 07 23:03:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

