

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070723\
 Data File : VX036459.D
 Acq On : 07 Jul 2023 14:40
 Operator : JC/MD
 Sample : 03552-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Jul 07 23:05:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	154089	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	263688	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	240049	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	108717	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	100663	45.574	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.140%
35) Dibromofluoromethane	5.397	113	82098	47.218	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.440%
50) Toluene-d8	8.653	98	320173	49.303	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.600%
62) 4-Bromofluorobenzene	11.079	95	112798	49.313	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.620%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070723\
Data File : VX036459.D
Acq On : 07 Jul 2023 14:40
Operator : JC/MD
Sample : 03552-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Jul 07 23:05:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

