

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070725\
 Data File : VX046893.D
 Acq On : 07 Jul 2025 09:13
 Operator : JC/MD
 Sample : VX0707MBL01
 Misc : 5.00g/10mL/100ul/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0707MBL01

Quant Time: Jul 08 04:42:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 08 04:37:37 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	262695	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	451036	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	411755	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	203879	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	169520	48.847	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.700%
35) Dibromofluoromethane	5.385	113	144425	47.172	ug/l	-0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.340%
50) Toluene-d8	8.647	98	540014	49.991	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.980%
62) 4-Bromofluorobenzene	11.079	95	209365	50.997	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070725\
Data File : VX046893.D
Acq On : 07 Jul 2025 09:13
Operator : JC/MD
Sample : VX0707MBL01
Misc : 5.00g/10mL/100ul/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0707MBL01

Quant Time: Jul 08 04:42:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 08 04:37:37 2025
Response via : Initial Calibration

